

CERTIFICATE

of Product Conformity (QAL1)

Certificate No: 0000001016_05

Certified AMS: FMD 09 for Waste gas velocity

Manufacturer: Dr. Födisch Umweltmesstechnik AG
Zwenkauer Str. 159
04420 Markranstädt
Germany

Test Institute: TÜV Rheinland Energy GmbH

**This is to certify that the AMS has been tested
and found to comply with the standards
EN 15267-1 (2009), EN 15267-2 (2009), EN 15267-3 (2007),
EN 16911 (2013) and EN 14181 (2015).**

Certification is awarded in respect of the conditions stated in this certificate
(this certificate contains 7 pages).

The present certificate replaces certificate 0000001016_04 dated 28 July 2021.



Suitability Tested
EN 15267
QAL1 Certified
Regular
Surveillance

www.tuv.com
ID 0000001016

Publication in the German Federal Gazette
(BAnz) of 20 July 2012


German Environment Agency
Dessau, 29 July 2022

This certificate will expire on:
28 July 2027

TÜV Rheinland Energy GmbH
Cologne, 28 July 2022



Dr. Marcel Langner
Head of Section II 4.1



ppa. Dr. Peter Wilbring

www.umwelt-tuv.eu
tre@umwelt-tuv.eu
Tel. + 49 221 806-5200

TÜV Rheinland Energy GmbH
Am Grauen Stein
51105 Köln

Test institute accredited to EN ISO/IEC 17025 by DAkS (German Accreditation Body).
This accreditation is limited to the accreditation scope defined in the enclosure to the certificate D-PL-11120-02-00.

Test report:	936/21212361/C dated 20 March 2012
Initial certification:	19 August 2011
Expiry date:	28 July 2027
Certificate:	Renewal (of previous certificate 0000001016_04 dated 28. Juli 2021 valid until 28 July 2022)
Publication:	BAnz AT 20.07.2012 B11, Chap. II No. 2.2

Approved application

The tested AMS is suitable for use at plants according to Directive 2010/75/EU, chapter III (13th BImSchV:2009), chapter IV (17th BImSchV:2009), 30th BImSchV:2009, Directive 2015/2193/EC (44th BImSchV:2021), TA Luft:2002 as well as at plants according to the 27th BImSchV:1997. The measured ranges have been selected so as to ensure as broad a field of application as possible.

The suitability of the AMS for this application was assessed on the basis of a laboratory test and a 6 month field test at a waste incineration plant.

The AMS is approved for an ambient temperature range of -20° to +50°C.

The notification of suitability of the AMS, performance testing and the uncertainty calculation have been effected on the basis of the regulations applicable at the time of testing. As changes in legal provisions are possible, any potential user should ensure that this AMS is suitable for monitoring the flue gas velocity relevant to the application.

Any potential user should ensure, in consultation with the manufacturer, that this AMS is suitable for the installation at which it will be installed.

Note:

The legal regulations mentioned do not correspond to the current state of legislation in every case. Each user should, if necessary, in consultation with the competent authority, ensure that this AMS meets the legal requirements for the intended use. In addition, it cannot be ruled out that legal regulations governing the use of a measuring device for emission monitoring may change during the lifetime of the certificate.

Basis of the certification

This certification is based on:

- Test report 936/21212361/C dated 20 March 2012 of TÜV Rheinland Energie und Umwelt GmbH
- Suitability announced by the German Federal Environment Agency (UBA) as the relevant body
- The ongoing surveillance of the product and the manufacturing process

Publication in the German Federal Gazette: BAnz AT 20.07.2012 B11, Chap. II No. 2.2,
Announcement by UBA dated 06 July 2012:

AMS designation:

FMD 09 for velocity

Manufacturer:

Dr. Födisch Umweltmesstechnik AG, Markranstädt

Field of application:

For plants requiring official approval and for plants according to the 27th BImSchV

Measuring ranges during the performance test:

Component	Certification range	Supplementary measurement ranges	Unit
Velocity	2 - 30	2 - 60-	m/s

Software version:

Main Version: 2.0

I/O Version: 1.1

Restriction:

The lower limit of the velocity measurement range is 2 m/s.

Notes:

1. The maintenance interval is three months.
2. After any malfunction of the filter resulting in high dust loads, the probe must be checked for contamination and cleaned if necessary.
3. As a pressure transmitter, the SMAR LD301 can be used in a range from 0 to 500 Pa or from 0 to 1000 Pa.
4. Supplementary test (additional measuring range) to the announcement of the Federal Environment Agency of 23 February 2012 (BAnz. p. 920, chapter II number 2.1).

Test report:

TÜV Rheinland Energie und Umwelt GmbH, Cologne

Report No.: 936/21212361/C dated 20 March 2012

Publication in the German Federal Gazette: BAnz AT 05.03.2013 B10, chapter. V
notification 25, Announcement by UBA dated 12 February 2013:

**25 Notification as regards Federal Environment Agency (UBA) notices
of 6 July 2012 (BAnz AT 20.07.2012 B11, chapter II number 2.2)**

The current software versions of the measuring system FMD 09 for velocity of the
company Dr. Födisch Umweltmesstechnik AG are:

Main Version: 2.07

I/O Version: 1.13

Statement issued by TÜV Rheinland Energie und Umwelt GmbH
dated 15 October 2012

Certified product

This certificate applies to automated measurement systems conforming to the following description:

The volumetric flow measurement FMD 09 is based on the determination of the differential pressure in flowing flue gas with help of a back pressure probe and a pressure sensor. The measuring system uses an in-situ method. The measured values from the pressure transmitter are transferred as 4 - 20 mA measuring signal to the evaluation electronics which are located in the measuring device.

The evaluation unit takes into account the differential pressure signal and waste gas boundary conditions as well as the cross-section of the duct. The stack temperature is continuously measured by a temperature sensor (PT100) which is integrated in the back pressure probe. The flow signal can be corrected by the measured temperature in the evaluation electronic.

The output of the volumetric flow or speed signal is carried out by several freely assignable 4 - 20 mA analogue outputs whose measuring range can be varied. In addition, the exhaust gas temperature, for example, can be output via the analogue outputs. It is possible to show either the actual measurement value or a line chart on the instrument display.

The control and display unit is integrated into a weather protected housing. The display shows all measured values, the status information and parameters. Using a keyboard it is possible to configure the display and to adapt the parameters specific for the instrument.

Optionally it is possible to connect an absolute pressure transmitter, through which the absolute pressure at the measurement area can be determined. This was not installed in the version used for the performance test. The signal from the absolute pressure transmitter can be used for calculation in the DAHS. Offsetting by the evaluation electronics of the FMD 09 was not tested.

General notes

This certificate is based upon the equipment tested. The manufacturer is responsible for ensuring that on-going production complies with the requirements of the EN 15267. The manufacturer is required to maintain an approved quality management system controlling the manufacture of the certified product. Both the product and the quality management systems shall be subject to regular surveillance.

If a product of the current production does not conform to the certified product, TÜV Rheinland Energy GmbH must be notified at the address given on page 1.

A certification mark with an ID-Number that is specific to the certified product is presented on page 1 of this certificate. This certification mark may be applied to the product or used in advertising materials for the certified product.

This document as well as the certification mark remains property of TÜV Rheinland Energy GmbH. With revocation of the publication the certificate loses its validity. After the expiration of the certificate and on requests of the TÜV Rheinland Energy GmbH this document shall be returned and the certificate mark must not be employed anymore.

The relevant version of this certificate and its expiration is also accessible on the internet: gal1.de.

History of documents

Certification of FMD 09 is based on the documents listed below and the regular, continuous monitoring of the Quality Management System of the manufacturer:

Initial certification according to EN 15267

Certificate No. 0000001016_00: 19 August 2011
Expiry date of the certificate: 18 August 2016
Test report 936/21212361/A dated 23 March 2011
TÜV Rheinland Energie und Umwelt GmbH
Publication BAnz. 29 July 2011, No. 113, p. 2725, chapter II number 1.1
UBA announcement dated 15 July 2011

Supplementary testing according to EN 15267

Certificate No. 0000001016_01: 16 March 2012
Expiry date of the certificate: 25 July 2016
Test report 936/21212361/B dated 19 October 2011
TÜV Rheinland Energie und Umwelt GmbH
Publication BAnz. 02 March 2012, No. 36, p. 920, chapter II number 2.1
UBA announcement dated 23 February 2012

Supplementary testing according to EN 15267

Certificate No. 0000001016_02: 20 August 2012
Expiry date of the certificate: 28 July 2016
Test report 936/21212361/C dated 20 March 2012
TÜV Rheinland Energie und Umwelt GmbH
Publication BAnz AT 20.07.2012 B11, chapter II number 2.2
UBA announcement dated 6 July 2012

Notifications

Statement issued by TÜV Rheinland Energie und Umwelt GmbH dated 15 October 2012
Publication BAnz AT 05.03.2013 B10, chapter V notification 25
UBA announcement dated 12 February 2013
(Software changes)

Renewal of certificate

Certificate No. 0000001016_03: 22 July 2016
Expiry date of the certificate: 28 July 2021

Renewal of certificate

Certificate No. 0000001016_04: 28 July 2021
Expiry date of the certificate: 28 July 2022

Renewal of certificate

Certificate No. 0000001016_05: 29 July 2022
Expiry date of the certificate: 28 July 2027

Calculation of overall uncertainty according to EN 14181 and EN 15267-3

Measuring system

Manufacturer
AMS designation
Serial number of units under test
Measuring principle

Dr. Födisch Umweltmesstechnik AG
FMD 09
09130 / 09131
Differential pressure measurement

Test report

Test laboratory
Date of report

936/21212361/A / 936/21212361/B / 936/21212361/C
TÜV Rheinland Energie
2011-03-23 / 2011-10-19 / 2012-03-20

Measured component

Certification range

Velocity
2 - 30 m/s

Calculation of the combined standard uncertainty

Tested parameter

Standard deviation from paired measurements under field conditions *
Lack of fit
Zero drift from field test
Span drift from field test
Influence of ambient temperature at span
Influence of supply voltage

		u^2	
u_D	0.127 m/s	0.016	(m/s) ²
u_{lof}	-0.196 m/s	0.038	(m/s) ²
u_{dr}	0.000 m/s	0.000	(m/s) ²
u_{ds}	0.173 m/s	0.030	(m/s) ²
u_t	0.058 m/s	0.003	(m/s) ²
u_v	0.059 m/s	0.003	(m/s) ²

* The larger value is used :
"Repeatability standard deviation at set point" or
"Standard deviation from paired measurements under field conditions"

Combined standard uncertainty (u_c)
Total expanded uncertainty

$$u_c = \sqrt{\sum (u_{max, j})^2} \quad 0.30 \text{ m/s}$$

$$U = u_c * k = u_c * 1.96 \quad 0.59 \text{ m/s}$$

Relative total expanded uncertainty
Requirement of 2010/75/EU
Requirement of EN 15267-3

U in % of the range 30 m/s	2.0
U in % of the range 30 m/s	10.0 **
U in % of the range 30 m/s	7.5

** The EU-directive 2010/75/EC on industrial emissions does not define requirements for this component.
A value of 10.0 % was used instead.