

# CONFIRMATION

Notification: 0000026912\_04\_02  
on changes according to EN 15267 regarding certificate 0000026912\_04 dated 12 June 2019

---

**AMS designation:** BAM 1020 PM<sub>2,5</sub> mit PM<sub>2,5</sub>-Vorabscheider for PM<sub>2,5</sub>

**Manufacturer:** Met One Instruments Inc.  
1600, Washington Blvd.  
Grants Pass, Oregon 97526,  
USA

---

## German Federal Environmental Agency (UBA)

**Announcement regarding the uniform practice in  
monitoring emissions and ambient air  
dated 09 March 2022  
Publication: BAnz AT 11.04.2022 B10**

**VI. Notifications as regards the uniform practice in monitoring  
emissions and ambient air:**

**18 Notification as regards Federal Environment Agency (UBA) notices  
of 12 July 2010 (BAnz. p. 2597, chapter II number 1.1) and  
of 31 March 2021 (BAnz AT 03.05.2021 B9, chapter III notification 43)**

The current software version of the measuring device BAM-1020 for suspended particulate matter PM<sub>2,5</sub> of the company Met One Instruments, Inc. is: R9.3.0

In the future, the measuring device will basically have a touch-sensitive screen and a USB port will now be installed in the front panel.


On the rear panel there is now an electrical connection for a new generation of digital wheater sensors.


In this context, the previous temperature sensor (BX-592) will be replaced with the digital model BX-598, and the combined pressure and temperature sensor BX-596 will be replaced with the digital model BX-597A, which also measures humidity.

Statement issued by TÜV Rheinland Energy GmbH dated 08 December 2021

---

TÜV Rheinland Energy GmbH  
Cologne, 12 April 2022

  
i. V. Dipl.-Ing. Guido Baum

  
i. A. Dipl.-Ing. Carsten Röllig

[www.umwelt-tuv.eu](http://www.umwelt-tuv.eu)  
TRE@umwelt-tuv.eu  
Tel. +49 221 806-5200

TÜV Rheinland Energy GmbH  
Am Grauen Stein  
51105 Cologne

Test institute accredited to EN ISO/IEC 17025 by DAkkS (German Accreditation Body).  
This accreditation is limited to the accreditation scope defined in the enclosure to the certificate D-PL-11120-02-00.