

CONFIRMATION

Notification: 0000034863_03_01
on changes according to EN 15267 regarding certificate 0000034863_03 dated 05 November 2019

AMS designation: GC 5000 BTX version PID for benzene

Manufacturer: AMA Instruments GmbH
Lise-Meitner-Strasse 8
89081 Ulm
Germany

German Federal Environmental Agency (UBA)

**Announcement regarding the uniform practice in
monitoring emissions and ambient air
dated 28 June 2022
Publication: BAnz AT 28.07.2022 B4**

**III. Notifications as regards the uniform practice in monitoring
emissions and ambient air:**

**38 Notification as regards Federal Environment Agency (UBA) notices
of 28 June 2019 (BAnz AT 22.07.2019 B8, chapter III number 3.2)**

The current software versions of the measuring device gas chromatograph GC 5000
BTX version PID for benzene of AMA Instruments GmbH are:


Common.Control v. 1.1
SS.Control v. 2.1
AMA_PEAK.log v. 1.3


Various hardware changes have been made to the measuring equipment. These
concern the following points:

- new connector modules at the back of the device
- new electronic pressure controller
- optimization of valve drive 6-port valve
- new actuator for furnace damper
- minor housing adaptations

Statement issued by TÜV Rheinland Energy GmbH dated 26 April 2022

TÜV Rheinland Energy GmbH
Cologne, 01 August 2022


i. V. Dipl.-Ing. Guido Baum


i. A. Dipl.-Ing. Carsten Röllig

www.umwelt-tuv.eu
TRE@umwelt-tuv.eu
Tel. +49 221 806-5200

TÜV Rheinland Energy GmbH
Am Grauen Stein
51105 Cologne

Test institute accredited to EN ISO/IEC 17025 by DAkkS (German Accreditation Body).
This accreditation is limited to the accreditation scope defined in the enclosure to the certificate D-PL-11120-02-00.