

CONFIRMATION

Notification: 0000035011_02_03
on changes according to EN 15267 regarding certificate 0000035011_02 dated 05 November 2019

AMS designation: UmweltOffice (DAHS)

Manufacturer: Siempelkamp NIS Ingenieurgesellschaft mbH
Industriestraße 13
63755 Alzenau
Germany

German Federal Environmental Agency (UBA)

**Announcement regarding the uniform practice in
monitoring emissions and ambient air
dated 09 March 2022
Publication: BAnz AT 11.04.2022 B10**

**VI. Notifications as regards the uniform practice in monitoring
emissions and ambient air:**

**45 Notification as regards Federal Environment Agency (UBA) notices
of 28 June 2019 (BAnz AT 22.07.2019 B8, chapter IV number 1.3) and
of 31 March 2021 (BAnz AT 03.05.2021 B9, chapter III notification 44)**

The software for the DAHS UmweltOffice of the company Siempelkamp NIS Ingenieurgesellschaft mbH was extended by the evaluation according to the current 13th BImSchV (July 2021) and by the use of the alternative database PostgreSQL. The results are documented in the test report 936/21249876/B dated 6 January 2022.


The software also fulfills all requirements of the 44th BImSchV.


The current software versions for the emission evaluation system UmweltOffice / Talas (ID = 0000035011) of Siempelkamp NIS Ingenieurgesellschaft mbH are:

Data evaluation and parameterization: UmweltOffice: 7.4.0
Oracle database: 11.2, 11.2 XE, 12.2, 18c XE, 18c SE or 19c SE2
PostgreSQL: 13.3
Data acquisition: TALAS/net: 5.3 (000), TALAS/7 7.4 (000)

Statement issued by TÜV Rheinland Energy GmbH dated 06 January 2022

TÜV Rheinland Energy GmbH
Cologne, 12 April 2022


i. V. Dipl.-Ing. Guido Baum


i. A. Dipl.-Ing. Carsten Röllig

www.umwelt-tuv.eu
TRE@umwelt-tuv.eu
Tel. +49 221 806-5200

TÜV Rheinland Energy GmbH
Am Grauen Stein
51105 Cologne

Test institute accredited to EN ISO/IEC 17025 by DAkkS (German Accreditation Body).
This accreditation is limited to the accreditation scope defined in the enclosure to the certificate D-PL-11120-02-00.