

CERTIFICATE

of Product Conformity (QAL1)

Certificate No.: 0000039321_01

Certified AMS: MGA12 for CO, NO, SO₂, O₂ and CO₂

Manufacturer: Dr. Födisch Umweltmesstechnik AG
Zwenkauer Straße 159
04420 Markranstädt
Germany

Test Institute: TÜV Rheinland Energy GmbH

**This is to certify that the AMS has been tested
and found to comply with:**

**EN 15267-1: 2009, EN 15267-2: 2009, EN 15267-3: 2007
and EN 14181: 2004**

Certification is awarded in respect of the conditions stated in this certificate
(see also the following pages).

The present certificate replaces certificate 0000039321 of 29 April 2019.



Suitability Tested
EN 15267
QAL1 Certified
Regular
Surveillance

www.tuv.com
ID 0000039321

Publication in the German Federal Gazette
(BAnz.) of 1 April 2014

German Federal Environment Agency
Dessau, 1 April 2019


Dr. Marcel Langner
Head of Section II 4.1

This certificate will expire on:
30 June 2020

TÜV Rheinland Energy GmbH
Cologne, 31 March 2019


ppa. Dr. Peter Wilbring

www.umwelt-tuv.de / www.enviro-tuv.eu
tre@umwelt-tuv.de
Tel. +49 221 806-5200

TÜV Rheinland Energy GmbH
Am Grauen Stein
51105 Cologne

Accreditation according to EN ISO/IEC 17025:2018 and certified according to ISO 9001:2015.

Test report:	936/21219366/A of 19 September 2013
Initial certification:	01 April 2014
Expiry date:	30 June 2020
Publication:	BAnz AT 01 April 2014 B12, chapter I, No. 3.4

Approved application

The tested AMS is suitable for use at combustion plants according to Directive 2010/75/EU, chapter III and other plants requiring official approval. The tested ranges have been chosen with respect to the wide application range of the AMS.

The suitability of the AMS for this application was assessed on the basis of a laboratory test and a three-month field test at a lignite-fired power plant.

The AMS is approved for an ambient temperature range of +5 °C to +30 °C.

Any potential user should ensure, in consultation with the manufacturer, that this AMS is suitable for the installation at which it will be installed.

Basis of the certification

This certification is based on:

- test report 936/21219366/A of 19 September 2013 of TÜV Rheinland Energie und Umwelt GmbH
- suitability announced by the German Federal Environment Agency (UBA) as the relevant body
- the ongoing surveillance of the product and the manufacturing process
- publication in the German Federal Gazette (BAnz AT 01 April 2014 B12, chapter I, No. 3.4, Announcement by UBA from 27 February 2014)

AMS designation:

MGA12 for CO, NO, SO₂, O₂ and CO₂

Manufacturer:

Dr. Födisch Umweltmesstechnik AG, Markranstädt

Field of application:

For measurements at plants requiring official approval (Directive 2010/75/EU, chapter III combustion plants)

Measuring ranges during the performance test:

Components	Certification range	Supplementary range	Units
CO	0 - 125	0 - 1000	mg/m ³
NO	0 - 300	0 - 1000	mg/m ³
SO ₂	0 - 200	0 - 1000	mg/m ³
O ₂	0 - 25	-	Vol.-%
CO ₂	0 - 20	-	Vol.-%

Software version:

1.47

Restrictions:

1. The ambient temperature must not exceed +30 °C.
2. The performance criterion as related to the expanded uncertainty according to EN 15267-3 was not fulfilled for the component CO.

Note:

The maintenance interval is four weeks.

Test report:

TÜV Rheinland Energie und Umwelt GmbH, Cologne
Report No.: 936/21219366/A of 19 September 2013

Certified product

This certificate applies to automated measurement systems conforming to the following description:

The multi-component MGA12 AMS is a measuring system for continuous monitoring of CO, NO, SO₂, O₂ and CO₂ in exhaust gases.

The components CO, NO, SO₂ and CO₂ are monitored using infrared absorption; O₂ is measured with an electrochemical cell.

The tested AMS comprises the gas analyser which is positioned in a 19" rack housing. The analyser is located in a heated and ventilated system cabinet with the dimensions 2100 x 800 x 600 mm, which also houses the sample gas pump (MGP 12), the sample gas cooler (GCU 12), the connections for measurement values and signals, and other electronic parts for electricity supply. A pump supplies the sample gas cooler with 15 % concentration phosphoric acid in order to prevent SO₂ absorption.

The sample gas is fed to gas preparation via a heated sample gas probe (HSP 12) and a heated sample gas pipe (25 m). The sample gas probe is fitted with a ceramic filter which, like the sample gas pipe, is heated to 180 °C.

General notes

This certificate is based upon the equipment tested. The manufacturer is responsible for ensuring that on-going production complies with the requirements of the EN 15267. The manufacturer is required to maintain an approved quality management system controlling the manufacture of the certified product. Both the product and the quality management systems shall be subject to regular surveillance.

If a product of the current production does not conform to the certified product, TÜV Rheinland Energy GmbH must be notified at the address given on page 1.

A certification mark with an ID-Number that is specific to the certified product is presented on page 1 of this certificate. This can be applied to the product or used in publicity material for the certified product.

This document as well as the certification mark remains property of TÜV Rheinland Energy GmbH. With revocation of the publication the certificate loses its validity. After the expiration of the certificate and on requests of the TÜV Rheinland Energy GmbH this document shall be returned and the certificate mark must not be employed anymore.

The relevant version of this certificate and its expiration is also accessible on the internet: qal1.de.

Certification of MGA12 for CO, NO, SO₂, O₂ and CO₂ is based on the documents listed below and the regular, continuous monitoring of the Quality Management System of the manufacturer:

Initial certification according to EN 15267

Certificate No. 0000039321: 29 April 2014
Expiry date of the certificate: 31 March 2019

Test report: 936/21219366/A of 19 September 2013
TÜV Rheinland Energie und Umwelt GmbH, Cologne
Publication: BAnz AT 01 April 2014 B12, chapter I, No. 3.4
Announcement by UBA from 27 February 2014

Renewal of the certificate according to EN 15267

Certificate No. 0000039321_01: 1 April 2019
Expiry date of the certificate: 30 June 2020

Calculation of overall uncertainty according to EN 14181 and EN 15267-3

Measuring system

Manufacturer	Dr. Födisch Umweltmesstechnik AG
AMS designation	MGA 12
Serial number of units under test	12002 / 12003
Measuring principle	IR

Test report

Test laboratory	936/21219366/A TÜV Rheinland
Date of report	2013-09-19

Measured component

Certification range	CO 0 - 125 mg/m ³
---------------------	---------------------------------

Evaluation of the cross-sensitivity (CS)

(system with largest CS)

Sum of positive CS at zero point	0.00 mg/m ³
Sum of negative CS at zero point	0.00 mg/m ³
Sum of positive CS at span point	3.70 mg/m ³
Sum of negative CS at span point	-2.50 mg/m ³
Maximum sum of cross-sensitivities	3.70 mg/m ³
Uncertainty of cross-sensitivity	u_i 2.140 mg/m ³

Calculation of the combined standard uncertainty

Tested parameter

			u^2
Standard deviation from paired measurements under field conditions *	u_D 0.690 mg/m ³	0.476	(mg/m ³) ²
Lack of fit	u_{lof} -0.577 mg/m ³	0.333	(mg/m ³) ²
Zero drift from field test	$u_{d,z}$ -0.144 mg/m ³	0.021	(mg/m ³) ²
Span drift from field test	$u_{d,s}$ -1.588 mg/m ³	2.522	(mg/m ³) ²
Influence of ambient temperature at span	u_t 1.510 mg/m ³	2.280	(mg/m ³) ²
Influence of supply voltage	u_v 0.537 mg/m ³	0.288	(mg/m ³) ²
Cross-sensitivity (interference)	u_i 2.140 mg/m ³	4.580	(mg/m ³) ²
Influence of sample gas flow	u_p 0.346 mg/m ³	0.120	(mg/m ³) ²
Uncertainty of reference material at 70% of certification range	u_{rm} 1.010 mg/m ³	1.021	(mg/m ³) ²

* The larger value is used :

"Repeatability standard deviation at span" or

"Standard deviation from paired measurements under field conditions"

Combined standard uncertainty (u_c)	$u_c = \sqrt{\sum (u_{max,j})^2}$	3.41 mg/m ³
Total expanded uncertainty	$U = u_c * k = u_c * 1.96$	6.69 mg/m ³

Relative total expanded uncertainty

Requirement of 2010/75/EU	U in % of the ELV 80 mg/m³	8.4
Requirement of EN 15267-3	U in % of the ELV 80 mg/m³	10.0
	U in % of the ELV 80 mg/m³	7.5

Calculation of overall uncertainty according to EN 14181 and EN 15267-3

Measuring system

Manufacturer	Dr. Födisch Umweltmesstechnik AG
Name of measuring system	MGA 12
Serial number of the candidates	12002 / 12003
Measuring principle	IR

Test report

Test laboratory	TÜV Rheinland
Date of report	2013-09-19

Measured component

	NO
Certification range	0 - 250 mg/m ³

Evaluation of the cross sensitivity (CS)

(system with largest CS)

Sum of positive CS at zero point	0.00 mg/m ³
Sum of negative CS at zero point	0.00 mg/m ³
Sum of positive CS at reference point	6.30 mg/m ³
Sum of negative CS at reference point	0.00 mg/m ³
Maximum sum of cross sensitivities	6.30 mg/m ³
Uncertainty of cross sensitivity	3.637 mg/m ³

Calculation of the combined standard uncertainty

Tested parameter

			u ²	
Standard deviation from paired measurements under field conditions *	u _D	3.095 mg/m ³	9.579	(mg/m ³) ²
Lack of fit	u _{lof}	1.155 mg/m ³	1.334	(mg/m ³) ²
Zero drift from field test	u _{d,z}	3.320 mg/m ³	11.022	(mg/m ³) ²
Span drift from field test	u _{d,s}	3.753 mg/m ³	14.085	(mg/m ³) ²
Influence of ambient temperature at span	u _t	2.468 mg/m ³	6.091	(mg/m ³) ²
Influence of supply voltage	u _v	1.208 mg/m ³	1.459	(mg/m ³) ²
Cross sensitivity (interference)	u _i	3.640 mg/m ³	13.250	(mg/m ³) ²
Influence of sample gas flow	u _b	1.383 mg/m ³	1.913	(mg/m ³) ²
Uncertainty of reference material at 70% of certification range	u _{rm}	2.021 mg/m ³	4.083	(mg/m ³) ²

* The larger value is used :

"Repeatability standard deviation at span" or

"Standard deviation from paired measurements under field conditions"

Combined standard uncertainty (u _c)	$u_c = \sqrt{\sum (u_{max,j})^2}$	7.93 mg/m ³
Total expanded uncertainty	$U = u_c * k = u_c * 1.96$	15.53 mg/m ³

Relative total expanded uncertainty

Requirement of 2010/75/EU	U in % of the ELV 120 mg/m³	12.9
Requirement of EN 15267-3	U in % of the ELV 120 mg/m ³	20.0
	U in % of the ELV 120 mg/m ³	15.0

Calculation of overall uncertainty according to EN 14181 and EN 15267-3

Measuring system

Manufacturer	Dr. Födisch Umweltmesstechnik AG
AMS designation	MGA 12
Serial number of units under test	12002 / 12003
Measuring principle	IR

Test report

Test laboratory	936/21219366/A TÜV Rheinland
Date of report	2013-09-19

Measured component

	SO ₂
Certification range	0 - 200 mg/m ³

Evaluation of the cross-sensitivity (CS)

(system with largest CS)

Sum of positive CS at zero point	0.00 mg/m ³
Sum of negative CS at zero point	-2.64 mg/m ³
Sum of positive CS at span point	5.10 mg/m ³
Sum of negative CS at span point	-8.00 mg/m ³
Maximum sum of cross-sensitivities	-8.00 mg/m ³
Uncertainty of cross-sensitivity	u_i -4.619 mg/m ³

Calculation of the combined standard uncertainty

Tested parameter

			u^2
Standard deviation from paired measurements under field conditions *	u_D	3.291 mg/m ³	10.831 (mg/m ³) ²
Lack of fit	u_{lof}	1.155 mg/m ³	1.334 (mg/m ³) ²
Zero drift from field test	$u_{d,z}$	0.346 mg/m ³	0.120 (mg/m ³) ²
Span drift from field test	$u_{d,s}$	-2.656 mg/m ³	7.054 (mg/m ³) ²
Influence of ambient temperature at span	u_t	2.452 mg/m ³	6.012 (mg/m ³) ²
Influence of supply voltage	u_v	0.947 mg/m ³	0.897 (mg/m ³) ²
Cross-sensitivity (interference)	u_i	-4.619 mg/m ³	21.333 (mg/m ³) ²
Influence of sample gas flow	u_p	0.722 mg/m ³	0.521 (mg/m ³) ²
Uncertainty of reference material at 70% of certification range	u_{rm}	1.617 mg/m ³	2.613 (mg/m ³) ²

* The larger value is used :

"Repeatability standard deviation at span" or

"Standard deviation from paired measurements under field conditions"

Combined standard uncertainty (u_c)	$u_c = \sqrt{\sum (u_{max,j})^2}$	7.12 mg/m ³
Total expanded uncertainty	$U = u_c * k = u_c * 1.96$	13.96 mg/m ³

Relative total expanded uncertainty

Requirement of 2010/75/EU

Requirement of EN 15267-3

U in % of the ELV 130 mg/m³	10.7
U in % of the ELV 130 mg/m³	20.0
U in % of the ELV 130 mg/m ³	15.0

Calculation of overall uncertainty according to EN 14181 and EN 15267-3

Measuring system

Manufacturer	Dr. Födisch Umweltmesstechnik AG
AMS designation	MGA 12
Serial number of units under test	12002 / 12003
Measuring principle	electrochemical cell

Test report

Test laboratory	TÜV Rheinland
Date of report	2013-09-19

Measured component

Certification range	O ₂	0 - 25 Vol.-%
---------------------	----------------	---------------

Evaluation of the cross-sensitivity (CS)

(system with largest CS)

Sum of positive CS at zero point	0.00	Vol.-%
Sum of negative CS at zero point	0.00	Vol.-%
Sum of positive CS at span point	0.00	Vol.-%
Sum of negative CS at span point	0.00	Vol.-%
Maximum sum of cross-sensitivities	0.00	Vol.-%
Uncertainty of cross-sensitivity	u _i	0.000 Vol.-%

Calculation of the combined standard uncertainty

Tested parameter

				u ²
Standard deviation from paired measurements under field conditions *	u _D	0.091	Vol.-%	0.008 (Vol.-%) ²
Lack of fit	u _{lof}	0.014	Vol.-%	0.000 (Vol.-%) ²
Zero drift from field test	u _{d,z}	-0.064	Vol.-%	0.004 (Vol.-%) ²
Span drift from field test	u _{d,s}	-0.110	Vol.-%	0.012 (Vol.-%) ²
Influence of ambient temperature at span	u _t	0.070	Vol.-%	0.005 (Vol.-%) ²
Influence of supply voltage	u _v	0.059	Vol.-%	0.003 (Vol.-%) ²
Cross-sensitivity (interference)	u _i	0.000	Vol.-%	0.000 (Vol.-%) ²
Influence of sample gas flow	u _b	-0.018	Vol.-%	0.000 (Vol.-%) ²
Uncertainty of reference material at 70% of certification range	u _{rm}	0.202	Vol.-%	0.041 (Vol.-%) ²

* The larger value is used :

- "Repeatability standard deviation at span" or
- "Standard deviation from paired measurements under field conditions"

Combined standard uncertainty (u _c)	$u_c = \sqrt{\sum (u_{max,j})^2}$	0.27	Vol.-%
Total expanded uncertainty	$U = u_c * k = u_c * 1.96$	0.53	Vol.-%

Relative total expanded uncertainty

Requirement of 2010/75/EU	U in % of the range 25 Vol.-%	2.1
Requirement of EN 15267-3	U in % of the range 25 Vol.-%	10.0 **
	U in % of the range 25 Vol.-%	7.5

** For this component no requirements in the EC-directives 2010/75/EU on industrial emissions are given. The chosen value is recommended by the certification body.

Calculation of overall uncertainty according to EN 14181 and EN 15267-3

Measuring system

Manufacturer	Dr. Födisch Umweltmesstechnik AG
AMS designation	MGA 12
Serial number of units under test	12002 / 12003
Measuring principle	IR

Test report

Test laboratory	936/21219366/A
Date of report	TÜV Rheinland
	2013-09-19

Measured component

Certification range	CO ₂
	0 - 20 Vol.-%

Evaluation of the cross-sensitivity (CS)

(system with largest CS)

Sum of positive CS at zero point	0.00	Vol.-%
Sum of negative CS at zero point	0.00	Vol.-%
Sum of positive CS at span point	0.60	Vol.-%
Sum of negative CS at span point	-0.20	Vol.-%
Maximum sum of cross-sensitivities	0.60	Vol.-%
Uncertainty of cross-sensitivity	u_i	0.346 Vol.-%

Calculation of the combined standard uncertainty

Tested parameter

				u^2
Standard deviation from paired measurements under field conditions *	u_D	0.142	Vol.-%	0.020 (Vol.-%) ²
Lack of fit	u_{lof}	0.058	Vol.-%	0.003 (Vol.-%) ²
Zero drift from field test	$u_{d,z}$	-0.012	Vol.-%	0.000 (Vol.-%) ²
Span drift from field test	$u_{d,s}$	0.346	Vol.-%	0.120 (Vol.-%) ²
Influence of ambient temperature at span	u_t	0.252	Vol.-%	0.064 (Vol.-%) ²
Influence of supply voltage	u_v	0.060	Vol.-%	0.004 (Vol.-%) ²
Cross-sensitivity (interference)	u_i	0.346	Vol.-%	0.120 (Vol.-%) ²
Influence of sample gas flow	u_n	-0.041	Vol.-%	0.002 (Vol.-%) ²
Uncertainty of reference material at 70% of certification range	u_{rm}	0.162	Vol.-%	0.026 (Vol.-%) ²

* The larger value is used :

"Repeatability standard deviation at span" or

"Standard deviation from paired measurements under field conditions"

Combined standard uncertainty (u_c)

$$u_c = \sqrt{\sum (u_{max,j})^2} \quad 0.60 \text{ Vol.-%}$$

Total expanded uncertainty

$$U = u_c * k = u_c * 1.96 \quad 1.17 \text{ Vol.-%}$$

Relative total expanded uncertainty

U in % of the range 20 Vol.-% **5.9**

Requirement of 2010/75/EU

U in % of the range 20 Vol.-% **10.0**

Requirement of EN 15267-3

U in % of the range 20 Vol.-% **7.5**