



CERTIFICATE

of Product Conformity (QAL1)

Certificate No.: 0000040201_03

Certified AMS: D-R 320 for dust

Manufacturer: DURAG GmbH

Kollaustraße 105 22453 Hamburg Germany

Test Institute: TÜV Rheinland Energy GmbH

This is to certify that the AMS has been tested and found to comply with:

EN 15267-1: 2009, EN 15267-2: 2009, EN 15267-3: 2007 and EN 14181: 2004

Certification is awarded in respect of the conditions stated in this certificate (see also the following pages).

The present certificate replaces certificate no. 0000040201_02 of 30 September 2015



Suitability Tested EN 15267 QAL1 Certified Regular Surveillance

www.tuv.com

Publication in the German Federal Gazette (BAnz.) of 26 August 2015

German Federal Environment Agency Dessau, 1 April 2019

Dr. Marcel Langner Head of Section II 4.1 This certificate will expire on: 30 June 2020

TÜV Rheinland Energy GmbH Cologne, 31 March 2019

1. Pet 619

ppa. Dr. Peter Wilbring

www.umwelt-tuv.de / www.eco-tuv.com

teu@umwelt-tuv.de Tel. +49 221 806-5200 TÜV Rheinland Energy GmbH

Am Grauen Stein 51105 Cologne

Accreditation according to EN ISO/IEC 17025:2018 and certified according to ISO 9001:2015.



Certificate: 0000040201 03 / 1 April 2019



Test report: 936/21225028/B of 2 March 2015

Initial certification: 1 April 2014
Expiry date: 30 June 2020

Publication: BAnz AT 26 August 2015 B4, chapter I number 1.1

Approved application

The tested AMS is suitable for use at combustion plants according to Directive 2010/75/EU, chapter III, at waste incineration plants according to Directive 2010/75/EU, chapter IV and other plants requiring official approval. The tested ranges have been chosen with respect to the wide application range of the AMS.

The suitability of the AMS for this application was assessed on the basis of a laboratory test and a field test of more than twelve month at a municipal waste incinerator.

The AMS is approved for an ambient temperature range of -40 °C to +60 °C

The notification of suitability of the AMS, performance testing, and the uncertainty calculation have been effected on the basis of the regulations valid at the time of performance testing. As changes in legal regulations are possible, any potential user should ensure that this AMS is suitable for monitoring the limit value relevant to the application.

Any potential user should ensure, in consultation with the manufacturer, that this AMS is suitable for the installation at which it will be installed.

Basis of the certification

This certification is based on:

- test report 936/21225028/B of 2 March 2015 of TÜV Rheinland Energie und Umwelt GmbH
- suitability announced by the German Federal Environment Agency (UBA) as the relevant body
- the ongoing surveillance of the product and the manufacturing process
- publication in the German Federal Gazette (BAnz AT 26 August 2015 B4, chapter I number 1.1, Announcement by UBA from 22 July 2015)



Certificate: 0000040201_03 / 1 April 2019



Measuring system:

D-R 320 for dust

Manufacturer:

DURAG GmbH, Hamburg

Field of application:

For measurements at plants requiring official approval (e.g. Directive 2010/75/EU on industrial emissions, chapters III and IV)

Measuring ranges during performance testing:

Component	Certification range	Unit
Dust	0 - 7,5*	mg/m³
*		

*equivalent to 0 – 500 SL

Component	Supplementary ranges				Unit
Dust	0 – 1000	0 – 4000	0 – 20.000	0 – 100	SL

Software versions:

D-R 320:

01.02R000

D-ISC 100:

01.03R0000

D-ESI 100:

1.1.015

Restrictions:

None

Notes:

- 1. The maintenance interval is six months.
- 2. The AMS can be used with the D-ISC 100 evaluation unit, the D-TB 200 supply unit or the D-TB 100 supply unit.
- 3. The AMS may be supplied with purge air either by way of the D-TB 200 supply unit or an external purge air supply.
- 4. The D-ISC 100 universal control unit has a digital Modbus RTU interface and a Modbus TCP in accordance with VDI 4201 Sheets 1 and 3 (EIA-485, serial and TCP/IP, Ethernet).
- 5. The D-R 320 measuring system is fitted with the Modbus RTU digital interface in accordance with VDI 4201 Sheets 1 and 3 (EIA-485, serial).
- 6. When using the D-R 320 measuring system with the D-ISC 100 universal control unit, the Modbus interface of the D-R 320 measuring system cannot be used. Instead, the Modbus digital interface of the D-ISC 100 universal control unit is used.
- 7. When using the AMS without the D-ISC 100 evaluation unit, the AMS shall be operated by means of the D-ESI 100 software on a customary PC / notebook / tablet.
- 8. The performance criterion as related to the determination coefficient R² of the calibration function according to EN 15267-3 was not fulfilled.
- 9. Supplementary testing (extension of maintenance interval) to Federal Environment Agency announcement of 17 July 2014 (BAnz AT 5 August 2014 B11, chapter I number 1.1 and chapter IV correction 1).

Test report:

TÜV Rheinland Energie und Umwelt GmbH, Cologne

Report no.: 936/21225028/B of 2 March 2015



Certificate: 0000040201_03 / 1 April 2019



Certified product

This certificate applies to automated measurement systems conforming to the following description:

The D-R 320 measuring system uses the principle of optical light scattering (backscattering) to measure dust. Measurements are made contact-free, continuous and without sampling in the flue gas flow above dewpoint. The red light from a laser diode is sent into the flue gas duct and illuminates the dust particles in the measuring volume. The light is scattered by the dust particles in the measuring volume and the light scattered back is detected by a photodiode. The proportion of the measured intensity of the scattered light to the intensity of the emitted light corresponds to the particle density in the measuring volume.

The measuring system consists of the following components:

- D-R 320 M measuring head and
- D-TB 100 electrical connection box for power supply or
- D-TB 200 supply unit with integrated purge air blower or
- D-ISC 100 universal control unit

When using either of the connection boxes D-TB 100 or D-TB 200, the D-R 320 measuring system is operated via PC by means of the D-ESI 100 control software. The D-ISC 100 control unit allows for operation of the AMS without a PC and may also provide additional data outputs. When using the connection units D-TB 100 and D-ISC 100, the measuring system needs to be fitted with an external purge air supply, for instance compressed air class 1 in accordance with ISO 8573-1:2010.

The connection boxes are merely used for mains supply, signal transmission (without affecting the actual processing of measured values), and purge air supply (D-TB 200 only). The generating of measured values as well as all calculation processes relevant to measuring (incl. the analogue and digital generating of measurements) occur directly within the measuring head.

The measuring system is available in two different versions for narrower and wider measurement channels (variants "narrow" and "wide"). With respect to the variant for narrow measurement channels, the measuring volume is situated at a distance ranging from 70 to 450 mm from the aperture. As far as the variant for wider measurement channels is concerned, the measuring volume is situated at a distance ranging from 240 to 1200 mm distance from the aperture / duct wall. The performance test was carried out with the variant for larger measurement channels.

Control measurements (control functions, zero point, contamination, span point) are made by use of an automatic swing-in "shuttle" (internal reference standard). Linearity checks can be performed by means of opacity filters that are placed in a measuring device which can be inserted in the measuring head. By swinging-in the internal reference standard device and dimming the light source, every settable measuring range (min. 0 to 100 SL) can be checked by means of this filter set. For this purpose it is not necessary to remove the instrument from the measuring location as it only needs to be opened up.

General notes

This certificate is based upon the equipment tested. The manufacturer is responsible for ensuring that on-going production complies with the requirements of the EN 15267. The manufacturer is required to maintain an approved quality management system controlling the manufacture of the certified product. Both the product and the quality management systems shall be subject to regular surveillance.

If a product of the current production does not conform to the certified product, TÜV Rheinland Energy GmbH must be notified at the address given on page 1.



Certificate: 0000040201 03 / 1 April 2019



A certification mark with an ID-Number that is specific to the certified product is presented on page 1 of this certificate. This can be applied to the product or used in publicity material for the certified product.

This document as well as the certification mark remains property of TÜV Rheinland Energy GmbH. With revocation of the publication the certificate loses its validity. After the expiration of the certificate and on requests of the TÜV Rheinland Energy GmbH this document shall be returned and the certificate mark must not be employed anymore.

The relevant version of this certificate and its expiration is also accessible on the internet: qal1.de.

Certification of D-R 320 for dust is based on the documents listed below and the regular, continuous monitoring of the Quality Management System of the manufacturer:

Initial certification according to EN 15267

Certificate No. 0000040201: 29 April 2014 Expiry date of the certificate: 31 March 2019

Test report: 936/21222219/A of 11 October 2013 TÜV Rheinland Energie und Umwelt GmbH, Cologne

Publication: BAnz AT 01 April 2014 B12, chapter I number 1.2

Announcement by UBA from 27 February 2014

Supplementary testing according to EN 15267

Certificate No. 0000040201_01: 9 September 2014 Expiry date of the certificate: 31 March 2019

Test report: 936/21222219/B of 2 April 2014

TÜV Rheinland Energie und Umwelt GmbH, Cologne

Publication: BAnz AT 5 August 2014 B11, chapter I number 1.1

UBA announcement of 17 July 2014

Supplementary testing according to EN 15267

Certificate No. 0000040201_02: 30 September 2015 Expiry date of the certificate: 31 March 2019

Test report: 936/21225028/B of 2 March 2015

TÜV Rheinland Energie und Umwelt GmbH, Cologne

Publication: BAnz AT 26 August 2015 B4, chapter I number 1.1

Announcement by UBA from 22 July 2015

Correction:

Publication: BAnz AT 5 August 2014 B11, chapter IV, no. 1 (correction report number) UBA announcement of 17 July 2014

Notification:

Statement of TÜV Rheinland Energie und Umwelt GmbH of 30 September 2014 Publication: BAnz AT 2 April 2015 B5, chapter IV notification 29 (new software version) UBA announcement of 25 February 2015

Renewal of the certificate according to EN 15267

Certificate No. 0000040201_03: 1 April 2019 Expiry date of the certificate: 30 June 2020



Certificate: 0000040201_03 / 1 April 2019

DURAG GmbH D-R 320

936/21225028/B TÜV Rheinland 2015-03-02

1235301 / 1235302 / 1236093 / 1236094 Scattered light analysis (back scattering)



Calculation of overall uncertainty according to EN 14181 and EN 15267-3

Measuring system
Manufacturer
AMS designation
Serial number of units under test
Measuring principle
Test report
Test laboratory
Date of report
Measured component
Certification range

Measured component	Dust	
Certification range	0 - 7.5	5 mg/m³

Calculation of the combined standard uncertainty					
Tested parameter				u ²	
Standard deviation from paired measurements under field conditions *	u_D	0.090	mg/m³	0.008	$(mg/m^3)^2$
Lack of fit	u _{lof}	-0.030	mg/m³	0.001	$(mg/m^3)^2$
Zero drift from field test	$u_{d,z}$	-0.078	mg/m³	0.006	$(mg/m^3)^2$
Span drift from field test	$u_{d,s}$	-0.095	mg/m³	0.009	$(mg/m^3)^2$
Influence of ambient temperature at span	u _t	0.020	mg/m³	0.000	$(mg/m^3)^2$
Influence of supply voltage	u_v	0.060	mg/m³	0.004	$(mg/m^3)^2$
Influence of sample gas pressure	u_p	0.000	mg/m³	0.000	$(mg/m^3)^2$
Uncertainty of reference material at 70% of certification range	u_{rm}	0.061	mg/m³	0.004	$(mg/m^3)^2$
* The larger value is used :					
"Repeatability standard deviation at span" or					

Standard deviation from paired measurements under field conditions			
Combined standard uncertainty (u _C) Total expanded uncertainty	$u_{c} = \sqrt{\sum_{k} (u_{\text{max } j})^{2}}$ $U = u_{c} * k = u_{c} * 1.96$	0.18 mg/ 0.35 mg/	

Relative total expanded uncertainty	U in % of the ELV 5 mg/m³	7.0
Requirement of 2010/75/EU	U in % of the ELV 5 mg/m³	30.0
Requirement of EN 15267-3	U in % of the ELV 5 mg/m³	22.5