

CERTIFICATE

of Product Conformity (QAL1)

Certificate No.: 0000040205_01

Certified AMS: Serinus 50 for SO₂

Manufacturer: Ecotech Pty Ltd.
1492 Ferntree Gully Road
Knoxfield, VIC, 3180
Australia

Test Institute: TÜV Rheinland Energy GmbH

**This is to certify that the AMS has been tested
and found to comply with:**

**VDI 4202-1: 2010, VDI 4203-3: 2010, EN 14212: 2012,
EN 15267-1: 2009 and EN 15267-2: 2009**

Certification is awarded in respect of the conditions stated in this certificate
(see also the following pages).

The present certificate replaces certificate 0000040205 of 29 April 2014.




Suitability Tested
Complying with
2008/50/EC
EN 15267
Regular
Surveillance
www.tuv.com
ID 0000040205


Publication in the German Federal Gazette
(BAnz.) of 01 April 2014

This certificate will expire on:
30 June 2020

German Federal Environment Agency
Dessau, 1 April 2019

TÜV Rheinland Energy GmbH
Cologne, 31 March 2019


Dr. Marcel Langner
Head of Section II 4.1


ppa. Dr. Peter Wilbring

www.umwelt-tuv.de / www.eco-tuv.com
teu@umwelt-tuv.de
Tel. +49 221 806-5200

TÜV Rheinland Energy GmbH
Am Grauen Stein
51105 Cologne

Accreditation according to EN ISO/IEC 17025:2018 and certified according to ISO 9001:2015.

Test report: 936/21221977/B of 08 October 2013

Initial certification: 01 April 2014

Date of expiry: 30 June 2020

Publication: BAuz AT 01 April 2014 B12, chapter IV, No. 3.1

Approved application

The tested AMS is suitable for the continuous measurement of concentrations of sulphur dioxide in ambient air (stationary operation).

The suitability of the AMS for this application was assessed on the basis of a laboratory test and a three-month field test.

The AMS is approved for a temperature range of 0 °C to +30 °C.

Any potential user should ensure, in consultation with the manufacturer, that this AMS is suitable for ambient air applications at which it will be installed.

Basis of the certification

This certification is based on:

- test report 936/21221977/B of 08 October 2013 of TÜV Rheinland Energie und Umwelt GmbH
- suitability announced by the German Federal Environment Agency (UBA) as the relevant body
- the ongoing surveillance of the product and the manufacturing process
- publication in the German Federal Gazette: BAuz AT 01 April 2014 B12, chapter IV, No. 3.1
Announcement by UBA from 27 February 2014

AMS designation:

Serinus 50 for SO₂

Manufacturer:

Ecotech Pty Ltd., Knoxfield, Australia

Field of application:

Continuous measurement of concentrations of sulphur dioxide in ambient air (stationary operation)

Measuring range during the performance test:

Component	Certification range	Unit
sulfur dioxide	0 - 1000	µg/m ³

Software version:

Firmware: 2.09.0005

Restrictions:

None

Notes:

1. The measuring system has to be operated in a lockable measuring cabinet or container.
2. The test report on the performance test is available online at www.qal1.de.

Test institute:

TÜV Rheinland Energie und Umwelt GmbH, Cologne

Report No.: 936/21221977/B of 8 October 2013

Certified product

This certificate applies to automated measurement systems conforming to the following description:

The Serinus 50 measuring system continuously monitors concentrations of sulphur dioxide by means of the ultraviolet fluorescence method. The instrument is designed for the continuous measuring of sulphur dioxide in ambient air.

The Serinus 50 measures SO₂ with the following components and techniques:

- Hydrocarbon kicker
- UV lamp
- fluorescence cell
- optical band-pass filter
- photomultiplier tube (PMT)

The SO₂ concentration is automatically corrected for gas temperature and pressure changes and referenced to 0 °C, 20 °C or 25 °C at 1 atmosphere. This allows the Serinus 50 to sample in the most useful range of SO₂ ambient measurement (25-500 ppb SO₂ in air.)

The measurement of sulphur dioxide is based on classical fluorescence spectroscopy principles. Sulphur dioxide (SO₂) exhibits a strong ultraviolet (UV) absorption spectrum between 200 and 240nm. When SO₂ absorbs UV from this wavelength, photon emission occurs (300-420nm). The amount of fluorescence emitted is directly proportional to the SO₂ concentration.

The Serinus 50 follows these principles and measurement techniques:

- Sample air is passed through a hydrocarbon kicker which removes hydrocarbons.
- UV energy from zinc discharge lamp is passed through a UV band pass filter are used to produce radiation at 214nm.
- The radiation is focused into the fluorescence cell where it is absorbed by the SO₂ molecules.
- The SO₂ molecules then emit photons (fluorescent light) uniformly in all directions.
- Wavelengths between 310-350nm, which are specific to SO₂, pass through a band pass filter where they reach the photomultiplier and record a signal.
- A reference detector monitors the emission from the zinc lamp and is used to correct for fluctuations in lamp intensity.

Exhaust air is scrubbed with a charcoal scrubber to eliminate hydrocarbons and SO₂. This air is then clean enough for use in the hydrocarbon kicker to remove hydrocarbons from the incoming sample air.

The Serinus 50 sulphur dioxide analyser consists of five main assemblies:

- The pneumatics to transfer sample and exhaust gas.
- The sensors for the measurement of SO₂ (optical cell) and other relevant parameters.
- The control system which encompasses all circuit boards which are used to control all sensors and pneumatic components.
- The power supply which supplies power for all the instrument processors.
- The communication module to access data.

Particulate Filter

The particulate filter is a Teflon 5 micron (μm) filter with a diameter of 47 mm. This filter eliminates all particles larger than 5 μm that could interfere with sample measurement.

Hydrocarbon Kicker

The hydrocarbon kicker removes interfering hydrocarbons from the sample air. This is achieved by using counter current exchange, where an air with a lower concentration of hydrocarbons moves in an opposite direction to air with a higher concentration. The high concentrations of hydrocarbons diffuse through a selective permeation membrane to the low concentration exhaust air and are removed. Increasing the flow of the low concentration air also increases the rate of diffusion.

Sample gas pump

Manufacturer: Thomas, type: 617CD22-194 C

During performance testing the above-mentioned sample gas pump was used in the laboratory as well as in the field test. As far as the models Serinus 10 (ozone), Serinus 30 (CO) and Serinus 50 (SO_2) are concerned, one pump can be operated with up to two analysers. However, for the Serinus 40 (NO_x) one sample gas pump per analyser is required.

General notes

This certificate is based upon the equipment tested. The manufacturer is responsible for ensuring that on-going production complies with the requirements of the EN 15267. The manufacturer is required to maintain an approved quality management system controlling the manufacture of the certified product. Both the product and the quality management systems shall be subject to regular surveillance.

If a product of the current production does not conform to the certified product, TÜV Rheinland Energy GmbH must be notified at the address given on page 1.

A certification mark with an ID-Number that is specific to the certified product is presented on page 1 of this certificate. This can be applied to the product or used in publicity material for the certified product is presented on page 1 of this certificate.

This document as well as the certification mark remains property of TÜV Rheinland Energy GmbH. With revocation of the publication the certificate loses its validity. After the expiration of the certificate and on requests of the TÜV Rheinland Energy GmbH this document shall be returned and the certificate mark must not be employed anymore.

The relevant version of this certificate and the validity is also accessible on the internet: qal1.de.

Certification of Serinus 50 sulphur dioxide Analyzer is based on the documents listed below and the regular, continuous monitoring of the Quality Management System of the manufacturer:

Initial certification according to EN 15267:

Certificate No. 0000040205: 29 April 2014
Validity of the certificate until: 31 March 2019

Test report: 936/21221977/B of 08 October 2013
TÜV Rheinland Energie und Umwelt GmbH, Köln
Publication: BAnz AT 01 April 2014 B12, chapter IV, No. 3.1
Announcement by UBA from 27 February 2014

Renewal of the certificate according to EN 15267:

Certificate No. 0000040205_01: 1 April 2019
Validity of the certificate until: 30 June 2020

Expanded uncertainty based on the results of the laboratory testing of Device 1

Measured component	Measuring device:	SO ₂	Serial-No.:	13-0096 (Device 1)	1h-limit value:	132	mmol/mol
Performance characteristic							
No.		Performance criterion	Result	Partial uncertainty	Partial uncertainty	Square of partial uncertainty	
1	Repeatability standard deviation at zero	≤ 1.0 nmol/mol	0.050	u _{t,z}	0.02	0.0002	
2	Repeatability standard deviation at 1h-limit value	≤ 3.0 nmol/mol	0.240	u _{t,h}	0.07	0.0050	
3	"lack of fit" at 1h-limit value	≤ 4.0% of measured value	2.620	u _{fh}	2.00	3.9868	
4	Sensitivity coefficient of sample gas pressure at 1h-limit value	≤ 2.0 nmol/mol/kPa	0.340	u _{gp}	2.70	7.2852	
5	Sensitivity coefficient of sample gas temperature at 1h-limit value	≤ 1.0 nmol/mol/K	0.050	u _{gt}	0.40	0.1609	
6	Sensitivity coefficient of surrounding temperature at 1h-limit value	≤ 1.0 nmol/mol/K	0.305	u _{st}	2.47	6.1146	
7	Sensitivity coefficient of electrical voltage at 1h-limit value	≤ 0.30 nmol/mol/V	0.027	u _v	0.25	0.0608	
8a	Interferent H ₂ O with 21 nmol/mol	≤ 10 nmol/mol (Zero)	0.010	u _{H2O}	2.25	5.0688	
8b	Interferent H ₂ S with 200 nmol/mol	≤ 10 nmol/mol (Span)	3.040	u _{int, pos}	1.600		
8c	Interferent NH ₃ with 200 nmol/mol	≤ 5.0 nmol/mol (Zero)	2.390				
8d	Interferent NO with 500 nmol/mol	≤ 5.0 nmol/mol (Span)	0.290				
8e	Interferent NO ₂ with 200 nmol/mol	≤ 5.0 nmol/mol (Zero)	-0.290	or	5.83	34.0086	
8f	Interferent m-Xylene with 1 µmol/mol	≤ 5.0 nmol/mol (Span)	1.080				
9	Averaging effect	≤ 7.0% of measured value	3.420	u _{int, neg}	3.050		
18	Difference sample/calibration port	≤ 1.0%	2.850		u _{av}	-2.23	4.9861
21	Uncertainty of test gas	≤ 3.0%	0.100	0.100	0.220	0.29	0.0843
			0.740	0.740	u _{usc}	1.32	1.7424
			1.250	1.250	u _{cg}		
			1.0	1.0	u _c	7.9689	nmol/mol
			0.000	0.000	u	15.9379	nmol/mol
			0.0843	0.0843	W	12.07	%
			1.7424	1.7424	W _{req}	15	cp

Certificate:
 0000040205_01 / 1 April 2019

Expanded uncertainty based on the results of the laboratory testing of Device 2

Measured component	Measuring device:	Serial-No.:	13-0097 (Device 2)
	SO ₂	1h-limit value:	132 nmol/mol
No.	Performance characteristic	Performance criterion	Result
1	Repeatability standard deviation at zero	$\leq 1.0 \text{ nmol/mol}$	0.000
2	Repeatability standard deviation at 1h-limit value	$\leq 3.0 \text{ nmol/mol}$	0.230
3	"lack of fit" at 1h-limit value	$\leq 4.0\%$ of measured value	1.590
4	Sensitivity coefficient of sample gas pressure at 1h-limit value	$\leq 2.0 \text{ nmol/mol/kPa}$	0.270
5	Sensitivity coefficient of sample gas temperature at 1h-limit value	$\leq 1.0 \text{ nmol/mol/K}$	0.030
6	Sensitivity coefficient of surrounding temperature at 1h-limit value	$\leq 1.0 \text{ nmol/mol/K}$	0.152
7	Sensitivity coefficient of electrical voltage at 1h-limit value	$\leq 0.30 \text{ nmol/mol/V}$	0.028
8a	Interferent H ₂ O with 21 nmol/mol	$\leq 10 \text{ nmol/mol (Zero)}$	-0.510
		$\leq 10 \text{ nmol/mol (Span)}$	3.060
8b	Interferent H ₂ S with 200 nmol/mol	$\leq 5.0 \text{ nmol/mol (Zero)}$	1.410
		$\leq 5.0 \text{ nmol/mol (Span)}$	2.210
8c	Interferent NH ₃ with 200 nmol/mol	$\leq 5.0 \text{ nmol/mol (Zero)}$	-0.310
		$\leq 5.0 \text{ nmol/mol (Span)}$	0.230
8d	Interferent NO with 500 nmol/mol	$\leq 5.0 \text{ nmol/mol (Zero)}$	3.670
		$\leq 5.0 \text{ nmol/mol (Span)}$	4.160
8e	Interferent NO ₂ with 200 nmol/mol	$\leq 5.0 \text{ nmol/mol (Zero)}$	1.000
		$\leq 5.0 \text{ nmol/mol (Span)}$	0.310
8f	Interferent m-Xylene with 1 µmol/mol	$\leq 10 \text{ nmol/mol (Zero)}$	0.860
		$\leq 10 \text{ nmol/mol (Span)}$	2.660
9	Averaging effect	$\leq 7.0\%$ of measured value	-2.620
18	Difference sample/calibration port	$\leq 1.0\%$	0.280
21	Uncertainty of test gas	$\leq 3.0\%$	2.000
Combined standard uncertainty		u _c	0.37
Expanded uncertainty		u	1.7424
Relative expanded uncertainty		w	0.1366
Maximum allowed expanded uncertainty		w _{ref}	15
		%	15

Expanded uncertainty based on the results of the laboratory and field testing of Device 1

Measured component:	Measuring device:	Serial No.:	13-0096 (Device 1)			
No.	Performance characteristic	Performance criterion	Result	Partial uncertainty	Square of partial uncertainty	1h-limit value: nmol/mol
1	Repeatability standard deviation at zero	≤ 1.0 nmol/mol	0.050	$u_{t,z}$	0.02	0.0002
2	Repeatability standard deviation at 1h-limit value	≤ 3.0 nmol/mol	0.240	$u_{t,fh}$	not considered, as $u_{t,fh} = 0.07 < u_{t,f}$	-
3	"lack of fit" at 1h-limit value	≤ 4.0% of measured value	2.620	$u_{t,h}$	2.00	3.9868
4	Sensitivity coefficient of sample gas pressure at 1h-limit value	≤ 2.0 nmol/mol/kPa	0.340	$u_{t,p}$	2.70	7.2852
5	Sensitivity coefficient of sample gas temperature at 1h-limit value	≤ 1.0 nmol/mol/K	0.050	$u_{t,t}$	0.40	0.1609
6	Sensitivity coefficient of surrounding temperature at 1h-limit value	≤ 1.0 nmol/mol/K	0.305	$u_{t,s}$	2.47	6.1146
7	Sensitivity coefficient of electrical voltage at 1h-limit value	≤ 0.30 nmol/mol/V	0.027	$u_{t,V}$	0.25	0.0608
8a	Interferent H ₂ O with 21 nmol/mol	≤ 10 nmol/mol (Zero)	0.010			
8b	Interferent H ₂ S with 200 nmol/mol	≤ 10 nmol/mol (Span)	3.040	$u_{t,H2S}$	2.25	5.0688
8c	Interferent NH ₃ with 200 nmol/mol	≤ 5.0 nmol/mol (Zero)	2.390	$u_{t,NH3}$		
8d	Interferent NO with 500 nmol/mol	≤ 5.0 nmol/mol (Span)	1.600			
8e	Interferent NO ₂ with 200 nmol/mol	≤ 5.0 nmol/mol (Zero)	0.290			
8f	Interferent m-Xylene with 1 μmol/mol	≤ 10 nmol/mol (Zero)	3.420			
9	Averaging effect	≤ 7.0% of measured value	2.850	$u_{t,avg}$	5.83	34.0086
10	Reproducibility standard deviation under field conditions	≤ 5.0% of average over 3 months	0.100			
11	Long term drift at zero level	≤ 4.0 nmol/mol	0.740	$u_{t,f}$	4.94	24.3720
12	Long term drift at span level	≤ 5.0% of max of certification range	0.940	$u_{t,l,z}$	-0.54	0.2945
13	Difference sample/calibration port	≤ 1.0%	3.810	$u_{t,l,h}$	2.90	8.4310
14	Uncertainty of test gas	≤ 3.0%	0.220	$u_{t,sc}$	0.29	0.0843
15			2.000	$u_{t,g}$	1.32	1.7424
16				Combined standard uncertainty	u_c	9.8283
17				Expanded uncertainty	U	19.6567
18				Relative expanded uncertainty	W	14.89
19				Maximum allowed expanded uncertainty	W_{req}	15

Expanded uncertainty based on the results of the laboratory and field testing of Device 2

Measured component	Ecotech Serinus 50	Serial-No.:	13-0097 (Device 2)			
No.	Performance characteristic	Performance criterion	Result	Partial uncertainty	Square of partial uncertainty	1h-limit value: nmol/mol
1	Repeatability standard deviation at zero	≤ 1.0 nmol/mol	0.000	$u_{r,z}$	0.00	0.0000
2	Repeatability standard deviation at 1h-limit value	≤ 3.0 nmol/mol	0.230	$u_{r,fh}$	not considered, as $u_{r,h} = 0.06 < u_{r,f}$	-
3	"lack of fit" at 1h-limit value	≤ 4.0% of measured value	1.590	$u_{l,fh}$	1.21	1.4683
4	Sensitivity coefficient of sample gas pressure at 1h-limit value	≤ 2.0 nmol/mol/kPa	0.270	u_{dp}	2.14	4.5625
5	Sensitivity coefficient of sample gas temperature at 1h-limit value	≤ 1.0 nmol/mol/K	0.030	u_{gt}	0.24	0.0587
6	Sensitivity coefficient of surrounding temperature at 1h-limit value	≤ 1.0 nmol/mol/K	0.152	u_{st}	1.24	1.5295
7	Sensitivity coefficient of electrical voltage at 1h-limit value	≤ 0.30 nmol/moV/V	0.028	u_V	0.26	0.0701
8a	Interferent H ₂ O with 21 nmol/mol	≤ 10 nmol/mol (Zero)	-0.510	u_{H2O}	2.11	4.4660
8b	Interferent H ₂ S with 200 nmol/mol	≤ 5.0 nmol/mol (Zero)	1.410	$u_{int,200}$	2.210	5.0000
8c	Interferent NH ₃ with 200 nmol/mol	≤ 5.0 nmol/mol (Zero)	-0.310	u_{NH3}	0.230	0.3670
8d	Interferent NO with 500 nmol/mol	≤ 5.0 nmol/mol (Zero)	0.230	$u_{int,500}$	4.160	5.48
8e	Interferent NO ₂ with 200 nmol/mol	≤ 5.0 nmol/mol (Span)	1.000	u_{NO2}	2.660	30.0628
8f	Interferent m-Xylene with 1 µmol/mol	≤ 10 nmol/mol (Span)	0.310	$u_{int,span}$	0.860	3.9868
9	Averaging effect	≤ 7.0% of measured value	-2.620	u_{av}	-2.00	3.9868
10	Reproducibility standard deviation under field conditions	≤ 5.0% of average over 3 months	3.740	u_{rf}	4.94	24.3720
11	Long term drift at zero level	≤ 4.0 nmol/mol	1.470	$u_{dl,z}$	0.85	0.7203
12	Long term drift at span level	≤ 5.0% of max. of certification range	3.540	$u_{dl,ih}$	2.70	7.2784
18	Difference sample/calibration point	≤ 1.0%	0.280	u_{asc}	0.37	0.1366
21	Uncertainty of test gas	≤ 3.0%	2.000	u_{sg}	1.32	1.7424
	Combined standard uncertainty			u_c	8.9696	nmol/mol
	Expanded uncertainty			U	17.9393	nmol/mol
	Relative expanded uncertainty			W	13.59	%
	Maximum allowed expanded uncertainty			W_{seq}	15	%