

# CERTIFICATE

## of Product Conformity (QAL1)

Certificate No.: 0000040329

**Certified AMS:** AF22M for SO<sub>2</sub>

**Manufacturer:** Environnement S.A.  
111 Boulevard Robespierre  
78304 Poissy Cedex  
France

**Test Institute:** TÜV Rheinland Energie und Umwelt GmbH

**This is to certify that the AMS has been tested  
and found to comply with:**

**VDI 4202-1: 2002, VDI 4203-2: 2004, EN 14212: 2012,  
EN 15267-1: 2009, EN 15267-2: 2009**

Certification is awarded in respect of the conditions stated in this certificate  
(see also the following pages).



Suitability Tested  
Complying with  
2008/50/EC  
EN 15267  
Regular  
Surveillance  
[www.tuv.com](http://www.tuv.com)  
ID 0000040329

Publication in the German Federal Gazette  
(BArz.) of 07 March 2008

This certificate will expire on:  
31 March 2019

German Federal Environment Agency  
Dessau, 29 April 2014

TÜV Rheinland Energie und Umwelt GmbH  
Cologne, 28 April 2014

i. A. Dr. Marcel Langner

ppa. Dr. Peter Wilbring

[www.umwelt-tuv.de](http://www.umwelt-tuv.de) / [www.eco-tuv.com](http://www.eco-tuv.com)  
teu@umwelt-tuv.de  
Tel. +49 221 806-5200

TÜV Rheinland Energie und Umwelt GmbH  
Am Grauen Stein  
51105 Cologne

Accreditation according to EN ISO/IEC 17025 and certified according to ISO 9001:2008.

**Certificate:**  
0000040329 / 29 April 2014

**Test report:** 936/21206773/A of 09 November 2007  
Addendum 936/21221709/B of 28 September 2013

**Initial certification:** 01 April 2014

**Date of expiry:** 31 March 2019

**Publication:** BAnz AT 01 April 2014 B12, chapter VI, notification 19

**Approved application**

The certified AMS is suitable for continuous monitoring of sulphur dioxide in ambient air.

The suitability of the AMS for this application was assessed on the basis of a laboratory test and a three-month field test.

The AMS is approved for a temperature range of 0 °C to +30 °C.

Any potential user should ensure, in consultation with the manufacturer, that this AMS is suitable for ambient air applications at which it will be installed.

**Basis of the certification**

This certification is based on:

- test report 936/21206773/A of 09 November 2007 of TÜV Rheinland Immissionsschutz und Energiesysteme GmbH and Addendum 936/21221709/B of 28 September 2013 of TÜV Rheinland Energie und Umwelt GmbH
- suitability announced by the German Federal Environment Agency (UBA) as the relevant body
- the on-going surveillance of the product and the manufacturing process
- publication in the German Federal Gazette (BAnz. 07 March 2008, No. 38, p. 901, chapter III, No. 1.1, UBA publication from 14 February 2008)
- publication in the German Federal Gazette (BAnz AT 01 April 2014 B12, chapter VI, notification 19, UBA publication from 27 February 2014)

**AMS designation:**

AF22M for SO<sub>2</sub>

**Manufacturer:**

Environnement S.A., Poissy Cedex, France

Distribution in Germany:

Ansyco GmbH, Karlsruhe

**Field of application:**

For continuous monitoring of sulphur dioxide in ambient air.

**Measuring ranges during the performance test:**

SO<sub>2</sub>    0 - 700 µg/m<sup>3</sup>  
          0 - 1000 µg/m<sup>3</sup>

**Software version:**

V1.22

**Testing institute:**

TÜV Rheinland Immissionsschutz und Energiesysteme GmbH, Cologne

TÜV Rheinland Group

Report No.: 936/21206773/A of 9 November 2007

**Notification of announcement by the German Federal Environment Agency dated 14th February 2008 (BAnz. p. 901, Chapter III Number 1.1)**

The measuring system AF22M for SO<sub>2</sub> manufactured by Environnement fulfils the requirements of EN 14212 (November 2012). Furthermore, the manufacturing process and quality management system of the measuring system AF22M for SO<sub>2</sub> fulfil the requirements of EN 15267.

The test report of the performance test with report number 936/21206773/A and an addendum as an integral part of to the test report with report number 936/21221709/B can be viewed on the internet at [www.qal1.de](http://www.qal1.de).

Statement by TÜV Rheinland Energie und Umwelt GmbH dated 28th September 2013

**Certified product**

This certificate applies to automated measurement systems conforming to the following description:

The sample is taken by a Teflon tube connected to the back of the monitor through a pump placed at the end of the circuit. A Teflon filter provides dust protection.

The sample to be analyzed is firstly filtered by an elimination device of aromatic hydrocarbon molecule. This device consists of two concentric tubes. The internal tube is made of a special polymer. The sample to be analyzed with aromatic HC molecules comes in the internal tube. Aromatic HC molecules are transferred by permeation to the external tube with effect that the transfer is done in the direction: more HC molecules in gas to few HC molecules in gas. The pump creates a vacuum in the external tube, the partial pressure of aromatic compounds decreases due to activated charcoal filtration and molecules are evacuated to the outside of the internal tube.

The sample to be analyzed, exempt from HC molecules, is directed to a reaction chamber in which it is irradiated by an ultraviolet radiation centered at 214 nm, the absorption wavelength of SO<sub>2</sub> molecules. A photodiode measures the ultraviolet radiation generated by the UV lamp, through a mirror. This measurement is used during signal processing in order to compensate for any variation of the UV energy.

Molecules restore a specific fluorescence in the ultraviolet, which is optically filtered between 300 and 400 nm at the outlet in order to eliminate some interfering gases. This fluorescence is visualized by the PM tube placed near the reaction chamber.

At the start of each „zero-ref“, a shutter is placed between the UV lamp and the reaction chamber inlet for 40 seconds. This electrical zero corresponds to the PM tube darkness current and the offset voltage of the preamplifier, incorporated into the signal processing, it eliminates the possibility of drifts with temperature and time.

Then PM tube signal is amplified and is converted into digital values for processing by a microprocessor that calculates the average of measurement values, checks the alarms and carries out monitor operation diagnosis. These various values and information are displayed on an alphanumeric display unit on the monitor front panel.

The analyser AF22M measures sulphur dioxide (SO<sub>2</sub>) in ambient air. The measuring principle is based on UV fluorescence.

The intensity of the radiation absorbed by the sulphur dioxide in the interior of an optical chamber with length L follows the principle of the Beer-Lambert law:

$$i_a = i_0 \times (1 - e^{-\alpha L c})$$

whereby "I<sub>0</sub>" is the intensity at the entry to the chamber, "α" the characteristic absorption efficiency for SO<sub>2</sub> and "c" = [SO<sub>2</sub>], the concentration of the gas to be analysed.

The probability that an excited molecule fluoresces is also expressed by the following formula:

$$\frac{Kf}{Kf + Kq + Kd}$$

The intensity of the fluorescence received by the photomultiplier (PM) is thus expressed as follows:

$$i_f = Gi_a \frac{Kf}{Kf + Kq + Kd}$$

whereby G is a constant which is dependent on the illuminated proportion of the chamber measured by the PM. Thus:

$$i_f = Gi_0 \frac{Kf}{Kf + Kq + Kd} \times (1 - e^{-\alpha Lc})$$

In this case,  $\alpha Lc \ll 1$  and  $1 - e^{-\alpha Lc}$  can be developed in the first order as follows:

$$1 - e^{-\alpha Lc} \cong +\alpha Lc$$

The result is thus::

$$i_f = \frac{Gi_0 Kf \alpha L}{Kf + Kq + Kd} c = \beta \cdot c$$

The radiation absorbed by the PM is thus directly proportional to SO<sub>2</sub> concentration.

The measuring principle complies with the standard reference method as stipulated in EN 14212.

**General notes**

This certificate is based upon the equipment tested. The manufacturer is responsible for ensuring that on-going production complies with the requirements of the EN 15267. The manufacturer is required to maintain an approved quality management system controlling the manufacture of the certified product. Both the product and the quality management systems shall be subject to regular surveillance.

If a product of the current production does not conform to the certified product, TÜV Rheinland Energie und Umwelt GmbH must be notified at the address given on page 1.

A certification mark with an ID-Number that is specific to the certified product is presented on page 1 of this certificate. This can be applied to the product or used in publicity material for the certified product is presented on page 1 of this certificate.

This document as well as the certification mark remains property of TÜV Rheinland Energie und Umwelt GmbH. With revocation of the publication the certificate loses its validity. After the expiration of the certificate and on requests of the TÜV Rheinland Energie und Umwelt GmbH this document shall be returned and the certificate mark must not be employed anymore.

The relevant version of this certificate and the validity is also accessible on the internet: [qal1.de](http://qal1.de).

Certification of AF22M for SO<sub>2</sub> is based on the documents listed below and the regular, continuous monitoring of the Quality Management System of the manufacturer:

**Initial test:**

Test report: 936/21206773/A of 9 November 2007  
TÜV Rheinland Immissionsschutz und Energiesysteme GmbH, Cologne

Publication: BAnz. 07 March 2008, No. 38, p. 901, chapter III, No. 1.1  
Announcement by UBA from 14 February 2008

**Initial certification according to EN 15267:**

Certificate No. 0000040329: 29 April 2014

Expiration date of the certificate: 31 March 2019

Test report: 936/21206773/A of 9 November 2007  
TÜV Rheinland Immissionsschutz und Energiesysteme GmbH, Cologne

Addendum 936/21221709/B of 28 September 2013  
TÜV Rheinland Energie und Umwelt GmbH, Cologne

Publication: BAnz AT 01 April 2014 B12, chapter VI, notification 19  
Announcement by UBA from 27 February 2014

**Notification:**

Publication: BAnz AT 01 April 2014 B12, chapter VI, notification 19  
Announcement by UBA from 27 February 2014

### Calculation of overall uncertainty lab test (Device 1)

Measuring device:	Environnement AF22M	Measured component:	SO <sub>2</sub>	Serial-No.:	Gerät 1	1h-limit value:	132	nmol/mol
<b>No.</b>								
	<b>Performance characteristic</b>		<b>Performance criterion</b>	<b>Result</b>	<b>Partial uncertainty</b>		<b>Square of partial uncertainty</b>	
1	Repeatability standard deviation at zero		≤ 1.0 nmol/mol	0.000	U <sub>r,z</sub>	0.00	0.0000	
2	Repeatability standard deviation at 1h-limit value		≤ 3.0 nmol/mol	0.200	U <sub>r,h</sub>	0.05	0.0026	
3	"lack of fit" at 1h-limit value		≤ 4.0% of measured value	-1.500	U <sub>l,h</sub>	-1.14	1.3068	
4	Sensitivity coefficient of sample gas pressure at 1h-limit value		≤ 2.0 nmol/mol/kPa	-0.140	U <sub>gp</sub>	-1.07	1.1384	
5	Sensitivity coefficient of sample gas temperature at 1h-limit value		≤ 1.0 nmol/mol/K	-0.023	U <sub>gt</sub>	-0.18	0.0307	
6	Sensitivity coefficient of surrounding temperature at 1h-limit value		≤ 1.0 nmol/mol/K	0.310	U <sub>st</sub>	2.36	5.5815	
7	Sensitivity coefficient of electrical voltage at 1h-limit value		≤ 0.30 nmol/mol/V	-0.010	U <sub>v</sub>	-0.10	0.0094	
8a	Interferent H <sub>2</sub> O with 21 nmol/mol		≤ 10 nmol/mol (Zero)	-0.300	U <sub>H2O</sub>	-1.44	2.0624	
8b	Interferent H <sub>2</sub> S with 200 nmol/mol		≤ 10 nmol/mol (Span)	-1.900	U <sub>int, pos</sub>			
8c	Interferent NH <sub>3</sub> with 200 nmol/mol		≤ 5.0 nmol/mol (Zero)	0.300	U <sub>int, pos</sub>			
8d	Interferent NO with 500 nmol/mol		≤ 5.0 nmol/mol (Span)	0.300	U <sub>int, neg</sub>			
8e	Interferent NO <sub>2</sub> with 200 nmol/mol		≤ 5.0 nmol/mol (Zero)	0.000	U <sub>int, neg</sub>			
8f	Interferent m-Xylene with 1 µmol/mol		≤ 10 nmol/mol (Zero)	0.000	U <sub>int, neg</sub>			
9	Averaging effect		≤ 7.0% of measured value	1.800	U <sub>av</sub>	1.37	1.8818	
18	Difference sample/calibration port		≤ 1.0%	0.190	U <sub>s,c</sub>	0.25	0.0629	
21	Uncertainty of test gas		≤ 3.0%	2.000	U <sub>c,g</sub>	1.32	1.7424	
	Combined standard uncertainty		U <sub>c</sub>		3.9522		nmol/mol	
	Expanded uncertainty		U		7.9045		nmol/mol	
	Relative expanded uncertainty		W		5.99		%	
	Maximum allowed expanded uncertainty		W <sub>req</sub>		15		%	

**Calculation of overall uncertainty lab test (Device 2)**

Measured component:	Measuring device:	Environnement AF22M	SO <sub>2</sub>	Serial-No.:	Gerät 2	1h-limit value:	132	nmol/mol
<b>Performance characteristic</b>								
No.		Performance criterion	Result	Partial uncertainty		Square of partial uncertainty		
1	Repeatability standard deviation at zero	≤ 1.0 nmol/mol	0.000	u <sub>r,z</sub>	0.00	0.0000	0.0000	
2	Repeatability standard deviation at 1h-limit value	≤ 3.0 nmol/mol	0.200	u <sub>r,h</sub>	0.05	0.0027	0.0027	
3	"lack of fit" at 1h-limit value	≤ 4.0% of measured value	0.100	u <sub>l,h</sub>	0.08	0.0058	0.0058	
4	Sensitivity coefficient of sample gas pressure at 1h-limit value	≤ 2.0 nmol/mol/kPa	0.250	u <sub>gp</sub>	1.91	3.6300	3.6300	
5	Sensitivity coefficient of sample gas temperature at 1h-limit value	≤ 1.0 nmol/mol/K	-0.012	u <sub>gt</sub>	-0.09	0.0084	0.0084	
6	Sensitivity coefficient of surrounding temperature at 1h-limit value	≤ 1.0 nmol/mol/K	0.340	u <sub>st</sub>	2.59	6.7140	6.7140	
7	Sensitivity coefficient of electrical voltage at 1h-limit value	≤ 0.30 nmol/mol/V	0.010	u <sub>v</sub>	0.10	0.0094	0.0094	
8a	Interferent H <sub>2</sub> O with 24 nmol/mol	≤ 10 nmol/mol (Zero)	-0.100	u <sub>H2O</sub>	-1.74	3.0327	3.0327	
8b	Interferent H <sub>2</sub> S with 200 nmol/mol	≤ 10 nmol/mol (Span)	-2.300	u <sub>int, pos</sub>				
8c	Interferent NH <sub>3</sub> with 200 nmol/mol	≤ 5.0 nmol/mol (Zero)	0.000	u <sub>int, neg</sub>				
8d	Interferent NO with 500 nmol/mol	≤ 5.0 nmol/mol (Span)	1.200	u <sub>int, pos</sub>				
8e	Interferent NO <sub>2</sub> with 200 nmol/mol	≤ 5.0 nmol/mol (Zero)	0.000	u <sub>int, neg</sub>				
8f	Interferent m-Xylene with 1 µmol/mol	≤ 10 nmol/mol (Zero)	-0.100	u <sub>int, pos</sub>				
9	Averaging effect	≤ 7.0% of measured value	2.900	u <sub>av</sub>	2.21	4.8845	4.8845	
18	Difference sample/calibration port	≤ 1.0%	0.010	u <sub>usc</sub>	0.01	0.0002	0.0002	
21	Uncertainty of test gas	≤ 3.0%	2.000	u <sub>cg</sub>	1.32	1.7324	1.7324	
Combined standard uncertainty								
				u <sub>c</sub>		5.0784	5.0784	nmol/mol
				Expanded uncertainty	U	10.1568	10.1568	nmol/mol
				Relative expanded uncertainty	W	7.69	7.69	%
				Maximum allowed expanded uncertainty	W <sub>req</sub>	15	15	%

**Calculation of overall uncertainty lab and field test (Device 1)**

Measured component:	Measuring device:	Environnement AF22M	Serial-No.:	Gerät 1	1h-limit value: nmol/mol	132
	No.	Performance characteristic	Performance criterion	Result	Partial uncertainty	Square of partial uncertainty
1		Repeatability standard deviation at zero	≤ 1.0 nmol/mol	0.000	u <sub>r,z</sub> u <sub>r,h</sub>	0.00 0.0000
2		Repeatability standard deviation at 1h-limit value	≤ 3.0 nmol/mol	0.200	u <sub>r,h</sub>	not considered, as u <sub>r,h</sub> = 0.05 < u <sub>r,f</sub>
3		"lack of fit" at 1h-limit value	≤ 4.0% of measured value	-1.500	u <sub>l,h</sub>	-1.14 1.3068
4		Sensitivity coefficient of sample gas pressure at 1h-limit value	≤ 2.0 nmol/mol/kPa	-0.140	u <sub>gp</sub>	-1.07 1.1384
5		Sensitivity coefficient of sample gas temperature at 1h-limit value	≤ 1.0 nmol/mol/K	-0.023	u <sub>gt</sub>	-0.18 0.0307
6		Sensitivity coefficient of surrounding temperature at 1h-limit value	≤ 1.0 nmol/mol/K	0.310	u <sub>st</sub>	2.36 5.5815
7		Sensitivity coefficient of electrical voltage at 1h-limit value	≤ 0.30 nmol/mol/V	-0.010	u <sub>v</sub>	-0.10 0.0094
8a		Interferent H <sub>2</sub> O with 21 nmol/mol	≤ 10 nmol/mol (Zero)	-0.300		
8b		Interferent H <sub>2</sub> S with 200 nmol/mol	≤ 10 nmol/mol (Span)	-1.900	u <sub>H2O</sub>	-1.44 2.0624
8c		Interferent NH <sub>3</sub> with 200 nmol/mol	≤ 5.0 nmol/mol (Span)	0.300	u <sub>int, pos</sub>	
8d		Interferent NO with 500 nmol/mol	≤ 5.0 nmol/mol (Zero)	0.000		
8e		Interferent NO <sub>2</sub> with 200 nmol/mol	≤ 5.0 nmol/mol (Span)	-0.400		
8f		Interferent m-Xylene with 1 µmol/mol	≤ 5.0 nmol/mol (Span)	0.400	or	1.34 1.8013
9		Averaging effect	≤ 7.0% of measured value	1.800	u <sub>av</sub>	1.37 1.8318
10		Reproducibility standard deviation under field conditions	≤ 5.0% of average over 3 months	2.240	u <sub>f</sub>	2.96 8.7427
11		Long term drift at zero level	≤ 4.0 nmol/mol	-0.610	u <sub>l,z</sub>	-0.35 0.1240
12		Long term drift at span level	≤ 5.0% of max. of certification range	-0.940	u <sub>l,h</sub>	-0.72 0.5132
18		Difference sample/calibration port	≤ 1.0%	0.190	u <sub>sc</sub>	0.25 0.0629
21		Uncertainty of test gas	≤ 3.0%	2.000	u <sub>c9</sub>	1.32 1.7424
		Combined standard uncertainty		u <sub>c</sub>	4.9997 9.9995	nmol/mol nmol/mol
		Expanded uncertainty		U	9.9995 7.58	%
		Relative expanded uncertainty		W	7.58 15	%
		Maximum allowed expanded uncertainty		W <sub>req</sub>		

**Calculation of overall uncertainty lab and field test (Device 2)**

Measured component:	Environment AF22M	Serial-No.:	Gerät 2			
No.	Performance characteristic	Performance criterion	Result	Partial uncertainty	Square of partial uncertainty	1h-limit value: nmol/mol
1	Repeatability standard deviation at zero	≤ 1.0 nmol/mol	0.000	$U_{f,z}$	0.00	0.0000
2	Repeatability standard deviation at 1h-limit value	≤ 3.0 nmol/mol	0.200	$U_{f,ih}$	not considered, as $U_{f,ih} = 0.05 < U_{f,f}$	-
3	"lack of fit" at 1h-limit value	≤ 4.0% of measured value	0.100	$U_{f,ih}$	0.08	0.0058
4	Sensitivity coefficient of sample gas pressure at 1h-limit value	≤ 2.0 nmol/mol/kPa	0.250	$U_{g,p}$	1.91	3.6300
5	Sensitivity coefficient of sample gas temperature at 1h-limit value	≤ 1.0 nmol/mol/K	-0.012	$U_{g,t}$	-0.09	0.0084
6	Sensitivity coefficient of surrounding temperature at 1h-limit value	≤ 1.0 nmol/mol/K	0.340	$U_{s,t}$	2.59	6.7140
7	Sensitivity coefficient of electrical voltage at 1h-limit value	≤ 0.30 nmol/mol/V	0.010	$U_V$	0.10	0.0094
8a	Interferent H <sub>2</sub> O with 21 nmol/mol	≤ 10 nmol/mol (Zero)	-0.100			
8b	Interferent H <sub>2</sub> S with 200 nmol/mol	≤ 5.0 nmol/mol (Zero)	-2.300	$U_{H_2O}$	-1.74	3.0327
8c	Interferent NH <sub>3</sub> with 200 nmol/mol	≤ 5.0 nmol/mol (Zero)	1.200	$U_{NH_3, pos}$		
8d	Interferent NO with 500 nmol/mol	≤ 5.0 nmol/mol (Zero)	0.500			
8e	Interferent NO <sub>2</sub> with 200 nmol/mol	≤ 5.0 nmol/mol (Zero)	0.500	or	2.40	5.7600
8f	Interferent m-Xylene with 1 µmol/mol	≤ 10 nmol/mol (Zero)	-0.100			
9	Averaging effect	≤ 7.0% of measured value	2.900	$U_{av}$	2.21	4.8845
10	Reproducibility standard deviation under field conditions	≤ 5.0% of average over 3 months	2.240	$U_{f,f}$	2.96	8.7427
11	Long term drift at zero level	≤ 4.0 nmol/mol	-0.690	$U_{d,l,z}$	-0.40	0.1587
12	Long term drift at span level	≤ 5.0% of max. of certification range	2.460	$U_{d,l,ih}$	1.87	3.5148
18	Difference sample/calibration port	≤ 1.0%	0.010	$U_{isc}$	0.01	0.0002
21	Uncertainty of test gas	≤ 3.0%	2.000	$U_{cg}$	1.32	1.7424
		Combined standard uncertainty	$U_c$			6.1809 nmol/mol
		Expanded uncertainty	$U$			12.3618 nmol/mol
		Relative expanded uncertainty	$W$			9.37 %
		Maximum allowed expanded uncertainty	$W_{req}$			15 %

# CONFIRMATION

Notification on changes according to EN 15267 regarding  
certificate 0000040329 dated 29 April 2014

---

**Measuring system:** AF22M for SO<sub>2</sub>

**Manufacturer:** Environnement S.A.  
111 Boulevard Robespierre  
78304 Poissy Cedex  
France

---

## German Federal Environmental Agency (UBA)

Announcement about the uniform practice in  
monitoring emissions and ambient air.

22 July 2015

Federal Gazette: BAuz AT 26 August 2015 B4

- V Notifications to the uniform practice for the continuous monitoring of emission and ambient air:
- 50 Notification as regards Federal Environment Agency (UBA) notices of 14 February 2008 (Federal Gazette (BAuz.) p. 901, chapter III number 1.1) and of 27 February 2014 (Federal Gazette (BAuz.) AT 1 April 2014 B12, chapter VI 19<sup>th</sup> notification)

The current software version for the AF22M measuring system for SO<sub>2</sub>, manufactured by Environnement S.A., is:

v1.0.4 (Calculation process)  
v3.6.f (Display process)

For additional means of communication, the AMS was equipped with a USB and a TCP/IP interface.


A protective device was added to the UV lamp bracket.

---

Statement of TÜV Rheinland Energie und Umwelt GmbH of 14 March 2015

TÜV Rheinland Energie und Umwelt GmbH  
Cologne, 04 November 2015

i. A. Dipl.-Ing. Ruth Steinhagen-Pinnow

  
[www.umwelt-tuv.de](http://www.umwelt-tuv.de)  
teu@umwelt-tuv.de  
Tel. +49 221 806-5200

i. A. Dipl. Ing. Carsten Röllig

  
TÜV Rheinland Energie und Umwelt GmbH  
Am Grauen Stein  
51105 Cologne

Test institute accredited to EN ISO/IEC 17025:2005 by DAkkS (German Accreditation Body). This accreditation is limited to the accreditation scope defined in the enclosure to the certificate D-PL-11120-02-00.	
---	--