

# CONFIRMATION

Notification: 0000040334\_01\_04  
on changes according to EN 15267 regarding certificate 0000040334\_01 dated 05 November 2019

---

**AMS designation:** MEAC300 (DAHS)

**Manufacturer:** SICK AG  
Nimburger Straße 11  
79276 Reute  
Germany

---

**German Federal Environmental Agency (UBA)**

**Announcement regarding the uniform practice in  
monitoring emissions and ambient air  
dated 29 June 2021  
Publication: BAnz AT 05.08.2021 B5**

**IV. Notifications as regards the uniform practice in monitoring  
emissions and ambient air:**

**18 Notification as regards Federal Environment Agency (UBA) notices  
of 28 June 2019 (BAnz AT 22.07.2019 B8, chapter IV number 1.2) and  
of 31 March 2021 (BAnz AT 03.05.2021 B9, chapter III notification 54)**

The latest software version for the MEAC300 emission evaluation system  
from SICK AG is:  
4.1.34.27


The versions up to 4.1.34.26 are no longer to be used, as incorrect output may occur in  
the logs.


The new version must be used in systems with remote data transmission, since  
messages and time identifiers are sometimes not transmitted correctly in the older  
versions.

Statement issued by TÜV Rheinland Energy GmbH dated 03 May 2021

---

TÜV Rheinland Energy GmbH  
Cologne, 16 August 2021

  
i. V. Dipl.-Ing. Guido Baum

  
i. A. Dipl.-Ing. Carsten Röllig

[www.umwelt-tuv.eu](http://www.umwelt-tuv.eu)  
TRE@umwelt-tuv.eu  
Tel. +49 221 806-5200

TÜV Rheinland Energy GmbH  
Am Grauen Stein  
51105 Cologne

Test institute accredited to EN ISO/IEC 17025 by DAkkS (German Accreditation Body).  
This accreditation is limited to the accreditation scope defined in the enclosure to the certificate D-PL-11120-02-00.