

CONFIRMATION

Notification: 0000040335_03_01

on changes according to EN 15267 regarding certificate 0000040335_03 dated 03 September 2021

AMS designation:	CEMSelect OEM for CO, NO, NO ₂ , NO _x , SO ₂ , O ₂ and CO ₂
Manufacturer:	Bühler Technologies GmbH Harkortstrasse 29 40880 Ratingen Germany

German Federal Environmental Agency (UBA)

Announcement regarding the uniform practice in monitoring emissions and ambient air dated 21 February 2023 Publication: BAnz AT 20.03.2023 B6

- IV. Notifications as regards the uniform practice in monitoring emissions and ambient air:
- 10 Notification as regards Federal Environment Agency (UBA) notice of 29 June 2021 (BAnz AT 05.08.2021 B5, chapter I number 4.3)

The current software versions for the CEMSelect OEM measuring system for CO, NO, NO_2 , NO_x , SO_2 , CO_2 and O_2 from Bühler Technologies GmbH for the measuring module Ultramat 23/BA 5000 are:

Ultramat 23-7MB2355 4.02.09, Ultramat 23-7MB2357 4.02.09, Ultramat 23-7MB2358 4.02.09, Ultramat 6 4.8.8, Ultramat 6-2K 4.8.8, Oxymat 6 4.8.8, Ultramat/Oxymat 6 4.8.8,

SIEMENS SIMATIC Set CEM CERT 7MB1957 Rev. 2.0, SIPROCESS UV600-7MB2621, BCU 9150883_3.003, Gas module 9137582_3.002, UV module 9139736_3.005

A new, equivalent chopper motor is used in the measuring modules of the Ultramat 6 series. An alternative manufacturer has been qualified for the bandwidth filter used in the IR measuring modules of the Ultramat 23 and Ultramat 6 series for the determination of NO and SO_2 .

Statement issued by TÜV Rheinland Energy GmbH dated 08 September 2022

TÜV Rheinland Energy GmbH Cologne, 25 April 2023

um i. V. Mpl.-Ing. Guido Baum

retty i. A. Dipl.-Ing. Carsten Röllig

www.umwelt-tuv.eu	TÜV Rheinland Energy GmbH
TRE@umwelt-tuv.eu	Am Grauen Stein
Tel. +49 221 806-5200	51105 Cologne

Test institute accredited to EN ISO/IEC 17025 by DAkkS (German Accreditation Body). This accreditation is limited to the accreditation scope defined in the enclosure to the certificate D-PL-11120-02-00.

01 notification on changes to no. certificate 0000040335_03