

CONFIRMATION

Notification: 0000043107_02_04
on changes according to EN 15267 regarding certificate 0000043107_02 dated 12 June 2019

AMS designation: APDA-372 / APDA-372 E for PM₁₀ and PM_{2.5}

Manufacturer: HORIBA Europe GmbH
Hans-Mess-Str. 6
61440 Oberursel /Ts.
Germany

German Federal Environmental Agency (UBA)

**Announcement regarding the uniform practice in
monitoring emissions and ambient air
dated 09 March 2022
Publication: BAnz AT 11.04.2022 B10**

**VI. Notifications as regards the uniform practice in monitoring
emissions and ambient air:**

**16 Notification as regards Federal Environment Agency (UBA) notices
of 25 February 2015 (BAnz AT 02.04.2015 B5, chapter III number 3.1) and
of 31 March 2021 (BAnz AT 03.05.2021 B9, chapter III notification 34)**


The measuring devices APDA-372 or APDA-372 E for PM₁₀ and PM_{2.5} of the company HORIBA Europe GmbH can be equipped in the future alternatively with the LED SBM-40-SC of the manufacturer Luminus. As an alternative to the sampling pump of type 1420VDP BLDC of the company Thomas Pumps used so far, the 2-head diaphragm pump of type NMP830.1.2KPDC-B HP 24V of the company KNF can be used in the future.


The current software version is: 100525.0014.0001.0001.0011

In addition to this version number, the following intermediate versions are also valid:
100468.0014.0001.0011 and 100524.0014.0001.0001.0011

Statement issued by TÜV Rheinland Energy GmbH dated 08 December 2021

TÜV Rheinland Energy GmbH
Cologne, 12 April 2022


i. V. Dipl.-Ing. Guido Baum


i. A. Dipl.-Ing. Carsten Röllig

www.umwelt-tuv.eu
TRE@umwelt-tuv.eu
Tel. +49 221 806-5200

TÜV Rheinland Energy GmbH
Am Grauen Stein
51105 Cologne

Test institute accredited to EN ISO/IEC 17025 by DAkkS (German Accreditation Body).
This accreditation is limited to the accreditation scope defined in the enclosure to the certificate D-PL-11120-02-00.