

CONFIRMATION

Notification: 0000062064_00_03
on changes according to EN 15267 regarding certificate 0000062064_00 dated 12 June 2019

AMS designation: 49iQ for O₃

Manufacturer: Thermo Fisher Scientific
27, Forge Parkway
Franklin, MA 02038
USA

German Federal Environmental Agency (UBA)

**Announcement regarding the uniform practice in
monitoring emissions and ambient air
dated 21 February 2023
Publication: BAnz AT 20.03.2023 B6**

**IV. Notifications as regards the uniform practice in monitoring
emissions and ambient air:**

**84 Notification as regards Federal Environment Agency (UBA) notices
of 27 February 2019 (BAnz AT 26.03.2019, chapter III number 3.1) and
of 28 June 2022 (BAnz AT 28.07.2022 B4, chapter III notification 51)**

The current software version of the 49iQ ambient air measuring system for O₃ from
Thermo Fisher Scientific is:


01.06.14.34444


In addition, the versions 01.06.13.34155 and 01.06.14.34442 are available.

To increase operational reliability, a "footprint" has been corrected on the optional
analog IO board of the measuring system.

Statement issued by TÜV Rheinland Energy GmbH dated 15 September 2022

TÜV Rheinland Energy GmbH
Cologne, 25 April 2023


i. V. Dipl.-Ing. Guido Baum


i. A. Dipl.-Ing. Carsten Röllig

www.umwelt-tuv.eu
TRE@umwelt-tuv.eu
Tel. +49 221 806-5200

TÜV Rheinland Energy GmbH
Am Grauen Stein
51105 Cologne

Test institute accredited to EN ISO/IEC 17025 by DAkkS (German Accreditation Body).
This accreditation is limited to the accreditation scope defined in the enclosure to the certificate D-PL-11120-02-00.