

# CONFIRMATION

Notification: 0000062067\_00\_02  
on changes according to EN 15267 regarding certificate 0000062067\_00 dated 12 June 2019

---

**AMS designation:** 42iQ for NO, NO<sub>2</sub> and NO<sub>x</sub>

**Manufacturer:** Thermo Fisher Scientific  
27, Forge Parkway  
Franklin, MA 02038  
USA

---

## German Federal Environmental Agency (UBA)

**Announcement regarding the uniform practice in  
monitoring emissions and ambient air  
dated 21 February 2023  
Publication: BA<sub>n</sub>z AT 20.03.2023 B6**

### **IV. Notifications as regards the uniform practice in monitoring emissions and ambient air:**

#### **81 Notification as regards Federal Environment Agency (UBA) notices of 27 February 2019 (BA<sub>n</sub>z AT 26.03.2019, chapter III number 1.1) and of 29 June 2021 (BA<sub>n</sub>z AT 05.08.2021 B5, chapter IV notification 9)**

The current software version of the 42iQ ambient air measuring system for NO, NO<sub>2</sub> and NO<sub>x</sub> from Thermo Fisher Scientific is:  
01.06.14.34444

In addition, the versions 01.06.11.33737, 01.06.12.34057, 01.06.12.34061,  
01.06.13.34157 and 01.06.14.34442 are available.


Alternatively, the K.N.F PU4877-N811 pump can be used for the measuring system.


To increase operational reliability, a "footprint" has been corrected on the optional  
analog IO board of the measuring system.

Statement issued by TÜV Rheinland Energy GmbH dated 15 September 2022

---

TÜV Rheinland Energy GmbH  
Cologne, 25 April 2023

  
i. V. Dipl.-Ing. Guido Baum

  
i. A. Dipl.-Ing. Carsten Röllig

[www.umwelt-tuv.eu](http://www.umwelt-tuv.eu)  
TRE@umwelt-tuv.eu  
Tel. +49 221 806-5200

TÜV Rheinland Energy GmbH  
Am Grauen Stein  
51105 Cologne

Test institute accredited to EN ISO/IEC 17025 by DAkkS (German Accreditation Body).  
This accreditation is limited to the accreditation scope defined in the enclosure to the certificate D-PL-11120-02-00.