

# CONFIRMATION

Notification: 0000072196\_01\_02  
on changes according to EN 15267 regarding certificate 0000072196\_01 dated 02 June 2021

---

**AMS designation:** MGAprime Q for CO, NO, NO<sub>2</sub>, N<sub>2</sub>O, SO<sub>2</sub>, O<sub>2</sub> and CO<sub>2</sub>

**Manufacturer:** MRU GmbH  
Fuchshalde 8  
74172 Neckarsulm  
Germany

---

## German Federal Environmental Agency (UBA)

**Announcement regarding the uniform practice in  
monitoring emissions and ambient air  
dated 28 June 2022  
Publication: BA<sub>n</sub>z AT 28.07.2022 B4**

### III. Notifications as regards the uniform practice in monitoring emissions and ambient air:

#### 17 Notification as regards Federal Environment Agency (UBA) notices of 31 March 2021 (BA<sub>n</sub>z AT 03.05.2021 B9, chapter I number 4.1) and of 29 June 2021 (BA<sub>n</sub>z AT 05.08.2021 B5, chapter IV notification 37)

The current software version of the portable measuring device MGAprimeQ for CO, NO, NO<sub>2</sub>, SO<sub>2</sub>, N<sub>2</sub>O, CO<sub>2</sub> and O<sub>2</sub> of MRU GmbH is:  
V1.002.008.

Furthermore, the intermediate versions V1.001.111; V1.001.112; V1.002.000; V1.002.001; V1.002.002; V1.002.003; V1.002.004; V1.002.005; V1.002.006; V1.002.007 can be used.


Furthermore, the measuring device can be equipped with the NDIR main board version B, the printed circuit board of the gas cooler module has been optimized. Changes to avoid condensate alarms were implemented and a thermal switch for the oxygen sensor was added.


The measuring device can be equipped with 5-meter and 10-meter heating line in addition to the previously used 3-meter heating line.

Statement issued by TÜV Rheinland Energy GmbH dated 26 April 2022

---

TÜV Rheinland Energy GmbH  
Cologne, 01 August 2022

  
i. V. Dipl.-Ing. Guido Baum

  
i. A. Dipl.-Ing. Carsten Röllig

[www.umwelt-tuv.eu](http://www.umwelt-tuv.eu)  
TRE@umwelt-tuv.eu  
Tel. +49 221 806-5200

TÜV Rheinland Energy GmbH  
Am Grauen Stein  
51105 Cologne

Test institute accredited to EN ISO/IEC 17025 by DAkkS (German Accreditation Body).  
This accreditation is limited to the accreditation scope defined in the enclosure to the certificate D-PL-11120-02-00.