



CERTIFICATE

of Product Conformity (QAL1)

Certificate No.: 0000081147_01

Certified AMS:

PFM 20 for dust

Manufacturer:

Dr. Födisch Umweltmesstechnik AG

Zwenkauer Str. 159 04420 Markranstädt

Germany

Test Institute:

TÜV Rheinland Energy GmbH

This is to certify that the AMS has been tested and found to comply with the standards EN 15267-1 (2009), EN 15267-2 (2009), EN 15267-3 (2007) as well as EN 14181 (2014).

Certification is awarded in respect of the conditions stated in this certificate (this certificate contains 6 pages).

The present certificate replaces certificate 0000081147_00 dated 31 May 2022.



Suitability Tested EN 15267 QAL1 Certified Regular Surveillance

www.tuv.com ID 0000081147

Publication in the German Federal Gazette (BAnz) of 02 August 2023

German Environment Agency Dessau, 05 September 2023 This certificate will expire on: 01 August 2028

TÜV Rheinland Energy GmbH Cologne, 04 September 2023

Mal &

Dr. Marcel Langner Head of Section II 4 De Pet ain

ppa. Dr. Peter Wilbring

www.umwelt-tuv.eu tre@umwelt-tuv.eu

tre@umwelt-tuv.eu Tel. + 49 221 806-5200 TÜV Rheinland Energy GmbH Am Grauen Stein 51105 Köln

Test institute accredited to EN ISO/IEC 17025 by DAkkS (German Accreditation Body).

This accreditation is limited to the accreditation scope defined in the enclosure to the certificate D-PL-11120-02-00.



Certificate:

0000081147_01 / 05 September 2023



Test report:

936/21255410/A dated 09 February 2023

Initial certification:

11 April 2022

Expiry date:

01 August 2028

Publication:

BAnz AT 02.08.2023 B7, chapter I No. 1.1

Approved application

The tested AMS is suitable for use at plants according to Directive 2010/75/EC, chapter III (combustion plants / 13th BlmSchV:2021), chapter IV (waste incineration plants / 17th BlmSchV:2021), Directive 2015/2193/EC (44th BlmSchV:2022), 30th BlmSchV:2019, TA Luft:2021 and 27th BlmSchV:2013. The measured ranges have been selected so as to ensure as broad a field of application as possible.

The suitability of the AMS for this application was assessed on the basis of a laboratory test and a nine month field test at an industrial dry plant for the production of ceramic floor coverings.

The AMS is approved for an ambient temperature range of -20 ° to +50 °C.

The notification of suitability of the AMS, performance testing and the uncertainty calculation have been effected on the basis of the regulations applicable at the time of testing. As changes in legal provisions are possible, any potential user should ensure that this AMS is suitable for monitoring the emission limit values relevant to the application.

Any potential user should ensure, in consultation with the manufacturer, that this AMS is suitable for the installation at which it will be installed.

Note:

The legal regulations mentioned correspond to the current state of legislation during certification. Each user should, if necessary, in consultation with the competent authority, ensure that this AMS meets the legal requirements for the intended use. In addition, it cannot be ruled out that legal regulations governing the use of a measuring device for emission monitoring may change during the lifetime of the certificate.

Basis of the certification

This certification is based on:

- Test report 936/21255410/A dated 09 February 2023 of TÜV Rheinland Energy GmbH
- Suitability announced by the German Federal Environment Agency (UBA) as the relevant body
- The ongoing surveillance of the product and the manufacturing process



Certificate:

0000081147_01 / 05 September 2023



Publication in the German Federal Gazette: BAnz AT 02.08.2023 B7, chapter I No. 1.1, Announcement by UBA dated 05 July 2023:

AMS designation:

PFM 20 for dust

Manufacturer:

Dr. Födisch Umweltmesstechnik AG, Markranstädt

Field of application:

For plants requiring official approval and for plants according to the 27th BImSchV.

Measuring ranges during the performance test:

Component	Certification range	A	Unit		
Dust	0 - 7.5	0 - 15	0 - 30	0 - 250	mg/m³

The measuring range 0 - 30 mg/m³ complies in the fieldtest to approx. 0 - 7.5 mg/m³ dust.

Software version: v1.43

Restrictions:

- 1. At plants with fluctuating waste gas velocities, the measuring system requires the signal of a QAL1-certified and calibrated waste gas velocity measuring device to compensate for the influence of velocity.
- 2. The measuring system must not be operated behind electrostatic precipitators.
- 3. The measuring system may only be used in waste gases that are not saturated with water vapour.

Notes:

- 1. The maintenance interval is three months.
- 2. The dust concentration is measured in wet waste gas under operating conditions.
- 3. Supplementary test (extension of the maintenance interval) with regard to the announcement of the Federal Environment Agency (UBA) of 09 March 2022 (BAnz AT 11.04.2022 B10, chapter I number 2.1).

Test institute: TÜV Rheinland Energy GmbH, Cologne Report No.: 936/21255410/A dated 09 February 2023



Certificate: 0000081147_01 / 05 September 2023



Certified product

This certificate applies to automated measurement systems conforming to the following description:

Measurement with the PFM 20 is carried out using the triboelectric measuring principle. The test gas in the waste gas flow is detected with the aid of the probe rod. Due to the dust particles flowing around and bouncing off, a charge exchange takes place between them and the probe rod. The derived current results in a signal that is dependent on the mechanical and electrical properties of the dust. The signal of the measuring system is also dependent on the exhaust gas velocity of the medium to be monitored. The dust-proportional signal, which is generated by the electronics integrated in the unit, is the measurement of the dust content.

The AMS can compensate for the influence of the waste gas velocity on the measured signal. To do this, it requires the signal from a QAL1-certified and calibrated waste gas velocity measuring system installed in the same measuring section.

The PFM 20 dust monitor consists of an in-situ probe with probe head and probe rod. The probe rod has a high-temperature coating for insulation. It is surrounded by a sleeve and an insulating body and is thus electrically isolated from the housing. The signal module with the evaluation electronics is located in the probe head.

The PFM 20 measuring system tested here consists of:

- The PFM 20 measuring system with the current software and
- A cable connecting the probe with the electronics.

The PFM20_HID software is required to operate the PFM 20 measuring system with a PC.

A standard notebook PC is required to parameterise the measuring system and to display the measured results of the AMS. The data is transferred via a specific USB cable. The measuring probe is mounted on the flue to be measured using a flange with a Tri-Clamp quick-release fastener.

The LinTest PFM 20 test system is available for the annual AST of the PFM 20 measuring system. The signal generator can be used to perform linearity tests as well as zero and span point checks.



Certificate: 0000081147_01 / 05 September 2023



General notes

This certificate is based upon the equipment tested. The manufacturer is responsible for ensuring that on-going production complies with the requirements of the EN 15267. The manufacturer is required to maintain an approved quality management system controlling the manufacture of the certified product. Both the product and the quality management systems shall be subject to regular surveillance.

If a product of the current production does not conform to the certified product, TÜV Rheinland Energy GmbH must be notified at the address given on page 1.

A certification mark with an ID-Number that is specific to the certified product is presented on page 1 of this certificate. This certification mark may be applied to the product or used in advertising materials for the certified product.

This document as well as the certification mark remains property of TÜV Rheinland Energy GmbH. With revocation of the publication the certificate loses its validity. After the expiration of the certificate and on requests of the TÜV Rheinland Energy GmbH this document shall be returned and the certificate mark must not be employed anymore.

The relevant version of this certificate and its expiration is also accessible on the internet: **qal1.de**.

History of documents

Certification of PFM 20 is based on the documents listed below and the regular, continuous monitoring of the Quality Management System of the manufacturer:

Initial certification according to EN 15267

Certificate No. 0000081147_00: 31 May 2022 Expiry date of the certificate: 10 April 2027

Test report: 936/21249601/A dated 10 November 2021

TÜV Rheinland Energy GmbH

Publication: BAnz AT 11.04.2022 B10, chapter I number 2.1

UBA announcement dated 9 March 2022

Supplementary testing according to EN 15267

Certificate No. 0000081147_01: 05 September 2023 Expiry date of the certificate: 01 August 2028 Test report: 936/21255410/A dated 9 February 2023

TÜV Rheinland Energy GmbH

Publication: BAnz AT 02.08.2023 B7, chapter I number 1.1

UBA announcement dated 5 July 2023



Requirement of EN 15267-3

Certificate:



0000081147_01 / 05 September 2023

Calculation of overall uncertainty according to EN 14181 and EN 15267-3

Measuring system								
Manufacturer		Dr. Födisch Umweltmesstechnik AG						
AMS designation		PFM 20						
Serial number of units under test		20001 / 20002						
Measuring principle		Triboelektrisch						
31								
Test report		936/21255410A						
Test laboratory	TÜV Rheinland							
Date of report	2023-02-09							
Measured component	Dust							
Certification range	0 -	7.5	mg/m³					
Calculation of the combined standard uncertainty								
Tested parameter				U ²				
Standard deviation from paired measurements under field conditions *	u_D	0.064	mg/m³	0.004	$(mg/m^3)^2$			
Lack of fit	u_{lof}	0.057	mg/m³	0.003	$(mg/m^3)^2$			
Zero drift from field test	$u_{d,z}$	0.030	mg/m³	0.001	$(mg/m^3)^2$			
Span drift from field test	$u_{d,s}$	0.081	mg/m³	0.007	$(mg/m^3)^2$			
Influence of ambient temperature at span	u_t	0.153	mg/m³	0.023	$(mg/m^3)^2$			
Influence of supply voltage	u_v	0.067	mg/m³	0.004	$(mg/m^3)^2$			
Uncertainty of reference material at 70% of certification range	u _{rm}	0.061	mg/m³	0.004	$(mg/m^3)^2$			
* The larger value is used :								
"Repeatability standard deviation at set point" or								
"Standard deviation from paired measurements under field conditions"								
Combined standard uncertainty (u.)		$\sqrt{\sum (u_m)}$	<u>}</u> 2	0.00	/ 3			
Combined standard uncertainty (u _C)		* k = u		0.22	mg/m³			
Total expanded uncertainty	$U = U_c$	$\kappa = u_0$	1.90	0.42	mg/m³			
Relative total expanded uncertainty	U in % of the ELV 5 mg/m³			8.4				
Requirement of 2010/75/EU		U in % of the ELV 5 mg/m³						
Requirement of 2010/13/EU	UIII 7	o OI LIIE	LLV 3 IIIQ/III°		30.0			

U in % of the ELV 5 mg/m³

22.5